


7471B

LB104329

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 1.7233e-004

C= -5.7233e-002

Rho= 0.9998134

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	169	0.000	169				
0.2	0.200	0.173	-0.027	1338	0.0 %	1338				
2.5	2.500	2.521	0.021	14959	0.0 %	14959				
5.0	5.000	5.021	0.021	29469	0.0 %	29469				
7.5	7.500	7.625	0.125	44578	0.0 %	44578				
10.0	10.000	9.888	-0.112	57711	0.0 %	57711				

LB104329 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Jul 2019 13:29:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	169	-	0.0000	7471B	PPB	31 Jul 2019 13:34:20	S	Std
0.2 - 1	50-2	1338	-	0.2000	7471B	PPB	31 Jul 2019 13:36:38	S	Std
2.5 - 1	52-5	14959	-	2.5000	7471B	PPB	31 Jul 2019 13:38:56	S	Std
5.0 - 1	55-0	29469	-	5.0000	7471B	PPB	31 Jul 2019 13:41:14	S	Std
7.5 - 1	57-5	44578	-	7.5000	7471B	PPB	31 Jul 2019 13:43:31	S	Std
10.0 - 1	5(0)-0	57711	-	10.0000	7471B	PPB	31 Jul 2019 13:45:49	S	Std
ICV54 - 1	ICV54	23281	3.9548	-	7471B	PPB	31 Jul 2019 13:50:12	U	SMPL
ICB54 - 1	ICB54	100	-0.0400	-	7471B	PPB	31 Jul 2019 13:52:28	U	SMPL
CCV24 - 1	CCV24	29246	4.9827	-	7471B	PPB	31 Jul 2019 13:54:44	U	SMPL
CCB24 - 1	CCB24	102	-0.0397	-	7471B	PPB	31 Jul 2019 13:56:59	U	SMPL
CRA - 1	CRA	1324	0.1709	-	7471B	PPB	31 Jul 2019 13:59:16	U	SMPL
HighStd - 1	HighStd	57898	9.9203	-	7471B	PPB	31 Jul 2019 14:01:32	U	SMPL
ChkStd - 1	ChkStd	40897	6.9905	-	7471B	PPB	31 Jul 2019 14:03:48	U	SMPL
PB121871BL - 1	PBS	109	-0.0384	-	7471B	PPB	31 Jul 2019 14:06:06	U	SMPL
PB121871BS - 1	LCSS	21711	3.6842	-	7471B	PPB	31 Jul 2019 14:08:23	U	SMPL
K3618-09 - 1	HD-02-073019	1771	0.2480	-	7471B	PPB	31 Jul 2019 14:10:40	U	SMPL
K3618-09DUP - 1	HD-02-073019DUP	1573	0.2138	-	7471B	PPB	31 Jul 2019 14:12:57	U	SMPL
K3618-09MS - 1	HD-02-073019MS	23380	3.9718	-	7471B	PPB	31 Jul 2019 14:15:14	U	SMPL
K3618-09MSD - 1	HD-02-073019MSD	23421	3.9789	-	7471B	PPB	31 Jul 2019 14:17:32	U	SMPL
K4066-01 - 1	TP-1-A	121638	20.9046	-	7471B	PPB	31 Jul 2019 14:19:48	U	SMPL
CCV25 - 1	CCV25	28919	4.9264	-	7471B	PPB	31 Jul 2019 14:22:04	U	SMPL
CCB25 - 1	CCB25	77	-0.0440	-	7471B	PPB	31 Jul 2019 14:24:22	U	SMPL
K4066-07 - 1	TP-11	7380	1.2146	-	7471B	PPB	31 Jul 2019 14:26:38	U	SMPL
K4066-11 - 1	TP-12	4871	0.7822	-	7471B	PPB	31 Jul 2019 14:28:54	U	SMPL
K4066-15 - 1	TP-13	6385	1.0431	-	7471B	PPB	31 Jul 2019 14:31:11	U	SMPL
K4077-06 - 1	20190612-LOC-13	413	0.0139	-	7471B	PPB	31 Jul 2019 14:33:27	U	SMPL
K3618-09LX5 - 1		378	0.0079	-	7471B	PPB	31 Jul 2019 14:35:44	U	SMPL
K3618-09A - 1		23326	3.9625	-	7471B	PPB	31 Jul 2019 14:38:00	U	SMPL
K4066-01DLX5 - 1		24609	4.1836	-	7471B	PPB	31 Jul 2019 14:40:17	U	SMPL
CCV26 - 1	CCV26	28931	4.9284	-	7471B	PPB	31 Jul 2019 14:42:34	U	SMPL
CCB26 - 1	CCB26	87	-0.0422	-	7471B	PPB	31 Jul 2019 14:44:52	U	SMPL