



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104334

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/31/2019	12:50
2	K4066-18		1	9	100.00	>212.0	>212.0	07/31/2019	13:20
3	K4066-18DUP		1	9	100.00	>212.0	>212.0	07/31/2019	13:45
4	K4079-02		1	7	100.00	>212.0	>212.0	07/31/2019	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP73119 WorkList ID : 129247 Department : Wet-Chemistry Date : 07-31-2019 09:53:14

LB 104334

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method	
08/12/2019	Water	K4066-18	C	Flash Point	Cool 4 deg C	PSEG01	N11	TP-13	07/29/2019	1010A
08/02/2019	Water	K4079-02	D	Flash Point	Cool 4 deg C	DVIR01	O41	TANK-4	07/30/2019	1010A

Date/Time 07/31/19 12:20
 Received by: KS
 Relinquished by: CP

Date/Time 07/31/19 14:45
 Received by: CP
 Relinquished by: KS