

LB104335


ISM02.4

INSTRUMENT ID : CU1

Linear

 μ Abs.:

46341



A= 0.0000e+000

B= 2.1359e-004

C= -3.1920e-002

Rho= 0.9995219

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.016	-0.016	75	0.000	75					-
0.05	0.050										
0.20	0.200	0.195	-0.005	1062	0.0 %	1062					-3
2.50	2.500	2.528	0.028	11986	0.0 %	11986					1
5.00	5.000	4.904	-0.096	23107	0.0 %	23107					-2
7.50	7.500	7.723	0.223	36308	0.0 %	36308					3
10.0	10.000	9.866	-0.134	46341	0.0 %	46341					-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	75	-		0 ISM02.4	PPB	7/31/2019 15:08	Std
	0.2 S0.2	1062	-		0.2 ISM02.4	PPB	7/31/2019 15:11	Std
	2.5 S2.5	11986	-		2.5 ISM02.4	PPB	7/31/2019 15:13	Std
	5 S5.0	23107	-		5 ISM02.4	PPB	7/31/2019 15:15	Std
	7.5 S7.5	36308	-		7.5 ISM02.4	PPB	7/31/2019 15:18	Std
	10 S10.0	46341	-		10 ISM02.4	PPB	7/31/2019 15:20	Std
ICV	ICV	18521	3.924 -		ISM02.4	PPB	7/31/2019 15:24	SMPL
ICB	ICB	12	-0.0294 -		ISM02.4	PPB	7/31/2019 15:26	SMPL
CCV027	CCV027	22673	4.8108 -		ISM02.4	PPB	7/31/2019 15:28	SMPL
CCB027	CCB027	18	-0.0281 -		ISM02.4	PPB	7/31/2019 15:30	SMPL
PB121872BL	PBS872	55	-0.0202 -		ISM02.4	PPB	7/31/2019 15:33	SMPL
K3924-01	MA52F0	970	0.1753 -		ISM02.4	PPB	7/31/2019 15:35	SMPL
K3924-02	MA52F0D	971	0.1755 -		ISM02.4	PPB	7/31/2019 15:37	SMPL
K3924-03	MA52F0S	15347	3.2461 -		ISM02.4	PPB	7/31/2019 15:39	SMPL
PB121873BL	PBS873	25	-0.0266 -		ISM02.4	PPB	7/31/2019 15:42	SMPL
K3963-01	MBEP25	21977	4.6622 -		ISM02.4	PPB	7/31/2019 15:44	SMPL
K3963-02	MBEP26	25919	5.5042 -		ISM02.4	PPB	7/31/2019 15:46	SMPL
K3963-03	MBEP29	14527	3.0709 -		ISM02.4	PPB	7/31/2019 15:49	SMPL
K3963-04	MBEP33	23441	4.9749 -		ISM02.4	PPB	7/31/2019 15:55	SMPL
K3963-05	MBEP27	16713	3.5378 -		ISM02.4	PPB	7/31/2019 15:57	SMPL
K3963-06	MBEP28	8442	1.7712 -		ISM02.4	PPB	7/31/2019 15:59	SMPL
K3963-07	MBEP30	29334	6.2336 -		ISM02.4	PPB	7/31/2019 16:02	SMPL
K3963-08	MBEP30D	29123	6.1885 -		ISM02.4	PPB	7/31/2019 16:04	SMPL
K3963-09	MBEP30S	39194	8.3396 -		ISM02.4	PPB	7/31/2019 16:09	SMPL
K3963-10	MBEP03	27984	5.9452 -		ISM02.4	PPB	7/31/2019 16:12	SMPL
K3963-11	MBEP04	51455	10.9584 -		ISM02.4	PPB	7/31/2019 16:14	SMPL
K3963-12	MBEP05	21332	4.5244 -		ISM02.4	PPB	7/31/2019 16:16	SMPL
K3963-13	MBEP06	65396	13.9361 -		ISM02.4	PPB	7/31/2019 16:18	SMPL
K3963-14	MBEP07	6343	1.3229 -		ISM02.4	PPB	7/31/2019 16:21	SMPL
K3963-15	MBEP08	6171	1.2862 -		ISM02.4	PPB	7/31/2019 16:23	SMPL
CCV028	CCV028	22767	4.8309 -		ISM02.4	PPB	7/31/2019 16:25	SMPL
CCB028	CCB028	-4	-0.0328 -		ISM02.4	PPB	7/31/2019 16:28	SMPL
K3963-16	MBEP09	2911	0.5898 -		ISM02.4	PPB	7/31/2019 16:30	SMPL
K3963-17	MBEP10	2851	0.577 -		ISM02.4	PPB	7/31/2019 16:32	SMPL
K3963-18	MBENZ4	10441	2.1982 -		ISM02.4	PPB	7/31/2019 16:34	SMPL
K3963-19	MBENZ5	14340	3.031 -		ISM02.4	PPB	7/31/2019 16:37	SMPL
K3963-20	MBENZ6	46144	9.824 -		ISM02.4	PPB	7/31/2019 16:39	SMPL
K3963-21	MBENZ7	9114	1.9148 -		ISM02.4	PPB	7/31/2019 16:41	SMPL
K3963-22	MBENZ8	13267	2.8018 -		ISM02.4	PPB	7/31/2019 16:43	SMPL
K3963-11DLX5		9837	2.0692 -		ISM02.4	PPB	7/31/2019 16:46	SMPL
K3963-13DLX5		12560	2.6508 -		ISM02.4	PPB	7/31/2019 16:48	SMPL
CCV029	CCV029	22933	4.8664 -		ISM02.4	PPB	7/31/2019 16:50	SMPL
CCB029	CCB029	49	-0.0215 -		ISM02.4	PPB	7/31/2019 16:53	SMPL