



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB104350
Analysis Date: 08/01/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 08/01/2019
Extraction IN Time: 07:40
Extraction OUT Time: 08:20

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB104350BL	LB104350BL	WATER	1.3	1000	100	2.1055	2.1055	3.02	2.1055	2.1055	0.0000	0
2	LB104350BS	LB104350BS	WATER	1.3	1000	100	2.7413	2.7413	3.01	2.7601	2.7601	0.0188	18.8
3	LB104350BSD	LB104350BSD	WATER	1.3	1000	100	2.9103	2.9103	3.03	2.9293	2.9293	0.0190	19
4	K4066-17	TP-13	WATER	1.3	1000	100	2.9932	2.9932	3.04	2.9941	2.9941	0.0009	0.9
5	K4091-01	GRAB	WATER	1.6	1000	100	2.9650	2.9650	3.05	3.0362	3.0362	0.0712	71.2



QC Batch# LB104350 Test: TPH Analysis Date: 08/01/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2539
pH Paper 0-14	M4240
Sodium Sulfate	EP1977
1:1 HCL	WP74116
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	10.00 ML	WP74118
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:46</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:45</u>		
Bal Check Time:	<u>08:00</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:15</u>
		Out Time:	<u>09:45</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:21</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:45</u>		
Bal Check Time:	<u>12:25</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:50</u>
		Out Time:	<u>11:20</u>		

25104350

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K4091 WorkList ID : 129270 Department : Wet-Chemistry Date : 08-01-2019 07:33:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2019	Water	K4066-17- K	TPH	Conc H2SO4 1c	PSEG01	N11	TP-13	07/29/2019	1664A
08/07/2019	Water	K4091-01 ✓	TPH	Conc H2SO4 1c	ARAM01	O21	GRAB	07/31/2019	1664A

Date/Time 08-01-19 7:36
Received by: JA
Relinquished by: CB

Date/Time 08-01-19 9:15
Received by: CB
Relinquished by: JD