



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104357

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/01/2019	11:57
2	CAL2	1	Water	NA	NA	20.1	6.99	08/01/2019	11:58
3	CAL3	1	Water	NA	NA	20.1	10.02	08/01/2019	11:59
4	ICV	1	Water	NA	NA	20.1	7.03	08/01/2019	12:01
5	CCV1	1	Water	NA	NA	20.1	2.01	08/01/2019	12:01
6	K4115-01	1	Water	NA	NA	23.3	4.60	08/01/2019	12:05
7	K4115-01DUP	1	Water	NA	NA	23.2	4.59	08/01/2019	12:06
8	CCV2	1	Water	NA	NA	20.1	12.02	08/01/2019	12:10

WORKLIST(Hardcopy Internal Chain)

2B104357

WorkList Name : PH K4114 w

WorkList ID : 129300

Department : Wet-Chemistry

Date : 08-01-2019 11:33:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2019	Water	K4115-01	✓ pH	Cool 4 deg C	STER14	O22	MINI-CELL	07/30/2019	SM 4500-H B

Date/Time

8/11/19 11:35

Received by:

D.P.

Relinquished by:

SP

Date/Time

8/11/19 12:30

Received by:

SP

Relinquished by:

D.P.