



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104359

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/01/2019	10:45
2	K4104-01		1	9	100.00	>212.0	>212.0	08/01/2019	11:10
3	K4104-01DUP		1	9	100.00	>212.0	>212.0	08/01/2019	11:35
4	K4104-09		1	7	100.00	>212.0	>212.0	08/01/2019	11:55
5	K4104-16		1	8	100.00	>212.0	>212.0	08/01/2019	12:20
6	K4104-18		1	9	100.00	>212.0	>212.0	08/01/2019	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8119-1 WorkList ID : 129293 Department : Wet-Chemistry Date : 08-01-2019 08:46:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2019	Water	K4104-01	Flash Point	Cool 4 deg C	PSEG01	N12	COMP-1	07/31/2019	1010A
08/07/2019	Water	K4104-09	Flash Point	Cool 4 deg C	PSEG01	N12	30383	07/31/2019	1010A
08/07/2019	Water	K4104-16	Flash Point	Cool 4 deg C	PSEG01	N12	COMP-2	07/31/2019	1010A
08/07/2019	Water	K4104-18	Flash Point	Cool 4 deg C	PSEG01	N12	COMP-3-WATER	07/31/2019	1010A

Date/Time 08/01/19 10:25
 Received by: KS
 Relinquished by: JP

Date/Time 08/01/19 12:45
 Received by: JP
 Relinquished by: KS