



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104360

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/01/2019	09:20
2	K4104-33		1	8	65.00	149.0	149.0	08/01/2019	09:45
3	K4104-33DUP		1	8	65.00	149.0	149.0	08/01/2019	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8119 WorkList ID : 129292 Department : Wet-Chemistry Date : 08-01-2019 08:45:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2019	Solid	K4104-33	Flash Point	Cool 4 deg C	PSEG01	N12	30079	07/31/2019	1010A

Date/Time 08/01/19 08:50
 Received by: KS
 Relinquished by: JP

Date/Time 08/01/19 10:20
 Received by: JP
 Relinquished by: KS