



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB104375

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/01/2019	15:10
2	CAL2	1	Water	NA	NA	20.1	6.99	08/01/2019	15:11
3	CAL3	1	Water	NA	NA	20.1	10.03	08/01/2019	15:12
4	ICV	1	Water	NA	NA	20.1	7.00	08/01/2019	15:13
5	CCV1	1	Water	NA	NA	20.1	2.02	08/01/2019	15:15
6	K4114-01	1	Water	NA	NA	19.9	6.73	08/01/2019	15:18
7	K4114-01DUP	1	Water	NA	NA	19.8	6.74	08/01/2019	15:19
8	CCV2	1	Water	NA	NA	20.1	12.01	08/01/2019	15:25

RB 104375

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k4126 WorkList ID : 129311 Department : Wet-Chemistry Date : 08-01-2019 14:51:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/01/2019	Water	K4114-01- X	pH	Cool 4 deg C	PSEG01	N13	CS-TEST-PIT	08/01/2019	9040C

Date/Time 08-01-19 15:00
 Received by: JA
 Relinquished by: CO

Date/Time 08-01-19 16:00
 Received by: CS
 Relinquished by: JS