



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/2/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 16:15
In Date: 08/01/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 08/02/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104376

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K4093-01	PL-02-080119	1	1.15	9.67	8.81	89.9
K4107-01	T-1	2	1.13	9.73	8.24	82.7
K4107-02	T-1A	3	1.14	9.69	7.92	79.3
K4107-03	T-2	4	1.15	9.69	7.8	77.9
K4107-04	T-2A	5	1.15	9.54	7.7	78.1
K4112-01	HA-04	6	1.00	2.00	2.00	100.0
K4112-02	HA-05	7	1.00	2.00	2.00	100.0
K4112-03	HA-21	8	1.00	2.00	2.00	100.0
K4120-01	T1-1	9	1.13	9.51	7.72	78.6
K4120-02	T1-2	10	1.12	9.8	7.83	77.3
K4120-03	T1-3	11	1.12	9.97	8.2	80.0
K4120-04	T2-1	12	1.13	9.82	7.67	75.3
K4120-05	T2-2	13	1.14	9.96	8.18	79.8
K4120-06	T2-3	14	1.13	9.9	7.62	74.0
K4120-07	T3-1	15	1.13	9.98	7.89	76.4
K4120-08	T3-2	16	1.13	9.98	8.07	78.4
K4120-09	T3-3	17	1.12	9.51	7.52	76.3
K4120-10	SW-4	18	1.13	9.73	7.52	74.3
K4120-11	SW-1	19	1.13	9.94	7.67	74.2
K4120-12	SW-2	20	1.14	9.61	7.54	75.6
K4120-13	SW-3	21	1.14	9.97	8.77	86.4
K4121-01	DRUM-COMP	22	1.00	2.00	2.00	100.0
K4122-01	HA-13	23	1.00	2.00	2.00	100.0
K4122-02	HA-14	24	1.00	2.00	2.00	100.0
K4122-03	HA-15	25	1.00	2.00	2.00	100.0
K4122-04	HA-16	26	1.00	2.00	2.00	100.0
K4122-05	HA-26	27	1.00	2.00	2.00	100.0
K4122-06	HA-27	28	1.00	2.00	2.00	100.0
K4122-07	HA-35	29	1.00	2.00	2.00	100.0
K4122-08	HA-36	30	1.00	2.00	2.00	100.0
K4122-09	HA-38	31	1.00	2.00	2.00	100.0
K4122-10	HA-39	32	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$