



Analytical Summary Report

Analysis Method: 9038

ANALYST: AMANDEEP

Parameter: Sulfate

SUPERVISOR REVIEW BY: apatel

Run Number: LB104387

Reagent/Standard	Lot/Log #
sulfate CCV 20ppm	WP77030
sulfate ICV 20ppm	WP77031
conditioning reagent	WP74905
Sulfate LCS std, 20ppm	WP77032
Sulfate LOD std 1.0PPM	WP77033
Sulfate Calibration std, 15ppm	WP77024
Sulfate Calibration std, 25ppm	WP77026
Sulfate Calibration std, 0ppm	WP76952
Sulfate Calibration std, 10ppm	WP77023
Sulfate Calibration std, 20ppm	WP77025
Sulfate Calibration std, 30ppm	WP77027
Sulfate Calibration std, 40ppm	WP77029
Sulfate Calibration std, 5ppm	WP77022
Sulfate Calibration std, 35ppm	WP77028
Sulfate LOQ Std- 5ppm	WP77034

Intercept: -0.7477

Slope: 3.2249

Regression: 0.999379

Seq	Lab ID	True Val (mg/l)	DF	MATRIX	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	Water	0.192	0.338	0.146	08/02/2019	10:15
2	CAL2	5	1	Water	0.226	14.700	14.474	08/02/2019	10:21
3	CAL3	10	1	Water	0.265	31.800	31.535	08/02/2019	10:27
4	CAL4	15	1	Water	0.245	46.600	46.355	08/02/2019	10:33
5	CAL5	20	1	Water	0.215	62.400	62.185	08/02/2019	10:38
6	CAL6	25	1	Water	0.295	82.700	82.405	08/02/2019	10:42
7	CAL7	30	1	Water	0.295	98.500	98.205	08/02/2019	10:48
8	CAL8	35	1	Water	0.297	112.000	111.703	08/02/2019	10:55
9	CAL9	40	1	Water	0.256	127.000	126.744	08/02/2019	11:02



Analytical Summary Report

Analysis Method: 9038
Parameter: Sulfate
Run Number: LB104387

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	MATRIX	Initial Reading	Final Reading	Difference	Result (mg/L)	AnalDate	Anal Time
1	ICV			20	1	Water	0.300	60.500	60.200	18.899	08/02/2019	11:08
2	ICB				1	Water	0.199	0.329	0.130	0.272	08/02/2019	11:15
3	CCV1			20	1	Water	0.217	61.200	60.983	19.142	08/02/2019	11:55
4	CCB1				1	Water	0.170	0.299	0.129	0.272	08/02/2019	12:01
5	LB104387BL	50	50		1	Water	0.196	0.301	0.105	0.264	08/02/2019	12:07
6	LB104387BS	50	50	20	1	Water	0.210	59.400	59.190	18.586	08/02/2019	12:13
7	K3591-09	50	50		1	Water	0.193	2.490	2.297	0.944	08/02/2019	12:18
8	CCV2			20	1	Water	0.201	62.500	62.299	19.550	08/02/2019	12:25
9	CCB2				1	Water	0.216	0.370	0.154	0.280	08/02/2019	12:32