



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104393

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/02/2019	09:15
2	K4114-02		1	8	100.00	>212.0	>212.0	08/02/2019	09:40
3	K4126-01		1	7	35.00	95.0	95.0	08/02/2019	10:05
4	K4126-01DUP		1	7	36.00	96.8	96.8	08/02/2019	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB104393

WorkList Name : FP8219 WorkList ID : 129340 Department : Wet-Chemistry Date : 08-02-2019 08:44:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2019	Water	K4114-02	Flash Point	Cool 4 deg C	PSEG01	N13	CS-TEST-PIT	08/01/2019	1010A
08/15/2019	Water	K4126-01	Flash Point	Cool 4 deg C	WARR01	O22	26-WARD-UNK-LIQUID	08/01/2019	1010A

Date/Time 08/02/19 08:45
 Received by: JP
 Relinquished by: JP

Date/Time 08/02/19 10:40
 Received by: JP
 Relinquished by: KS