



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104407

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/02/2019	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	08/02/2019	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	08/02/2019	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	08/02/2019	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	08/02/2019	16:16
6	K4158-01	1	Solid	20.02	20	23.4	6.88	08/02/2019	16:19
7	K4158-01DUP	1	Solid	20.03	20	23.5	6.89	08/02/2019	16:20
8	CCV2	1	Water	NA	NA	20.1	12.02	08/02/2019	16:21

25104407

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4152 WorkList ID : 129364 Department : Wet-Chemistry Date : 08-02-2019 15:56:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/02/2019	Solid	K4158-01	pH	Cool 4 deg C	TULL02	O51	FK-PS-01-055	08/02/2019	9045D

Date/Time 08.02.19 16:00
Received by: JN
Relinquished by: JN

Date/Time 08.02.19 17:35
Received by: JN
Relinquished by: JN