



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104408

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/02/2019	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	08/02/2019	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	08/02/2019	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	08/02/2019	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	08/02/2019	16:25
6	K4142-01	1	Water	NA	NA	25.1	7.30	08/02/2019	16:26
7	K4142-01DUP	1	Water	NA	NA	25.2	7.30	08/02/2019	16:27
8	CCV2	1	Water	NA	NA	20.1	12.02	08/02/2019	16:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4142 WorkList ID : 129365 Department : Wet-Chemistry Date : 08-02-2019 15:56:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/02/2019	Water	K4142-01	pH	Cool 4 deg C	PSEG01	N21	399L-241L	08/02/2019	9040C

Date/Time 08.02.19 16.00
 Received by: Sh
 Relinquished by: Sh

Date/Time 08.02.19 17:20
 Received by: Sh
 Relinquished by: Sh