



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104428

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/05/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	08/05/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	08/05/2019	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	08/05/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	08/05/2019	17:06
6	K4162-03	1	Solid	20.02	20	23.4	7.32	08/05/2019	17:09
7	K4162-03DUP	1	Solid	20.03	20	23.3	7.33	08/05/2019	17:10
8	K4162-06	1	Solid	20.05	20	23.6	6.24	08/05/2019	17:11
9	K4164-03	1	Solid	20.02	20	23.0	8.83	08/05/2019	17:12
10	K4164-06	1	Solid	20.02	20	22.6	5.35	08/05/2019	17:14
11	CCV2	1	Water	NA	NA	20.1	12.02	08/05/2019	17:15

18104628

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4161 WorkList ID : 129414 Department : Wet-Chemistry Date : 08-05-2019 16:05:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2019	Solid	K4162-03-17	Corrosivity	Cool 4 deg C	PSEG01	N22	RT-3407	08/05/2019	9045D
08/05/2019	Solid	K4162-06-17	Corrosivity	Cool 4 deg C	PSEG01	N22	RBR-2000050	08/05/2019	9045D
08/05/2019	Solid	K4164-03-17	Corrosivity	Cool 4 deg C	PSEG01	N22	RBR-2000030	08/05/2019	9045D
08/05/2019	Solid	K4164-06-17	Corrosivity	Cool 4 deg C	PSEG01	N22	RBR-2000051	08/05/2019	9045D

Date/Time 08/05/19 16:10
 Received by: JH
 Relinquished by: CH

Date/Time 08/05/19 16:02
 Received by: CH
 Relinquished by: JH