

LB104442


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

42236



A= 0.0000e+000

B= 2.3740e-004

C= 3.9716e-002

Rho= 0.9997425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.049	0.049	39	0.000	39				
0.2	0.200	0.230	0.030	801	0.0 %	801				
2.5	2.500	2.516	0.016	10432	0.0 %	10432				
5.0	5.000	4.817	-0.183	20122	0.0 %	20122				
7.5	7.500	7.522	0.022	31518	0.0 %	31518				
10.0	10.000	10.066	0.066	42236	0.0 %	42236				

LB104442 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Aug 2019 13:55:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	39	-	0.0000	7470A	PPB	06 Aug 2019 14:17:36	S	Std
0.2 - 1	50-2	801	-	0.2000	7470A	PPB	06 Aug 2019 14:19:44	S	Std
2.5 - 1	52-5	10432	-	2.5000	7470A	PPB	06 Aug 2019 14:21:53	S	Std
5.0 - 1	55-0	20122	-	5.0000	7470A	PPB	06 Aug 2019 14:24:02	S	Std
7.5 - 1	57-5	31518	-	7.5000	7470A	PPB	06 Aug 2019 14:26:10	S	Std
10.0 - 1	510-0	42236	-	10.0000	7470A	PPB	06 Aug 2019 14:28:18	S	Std
ICV60 - 1	ICV60	17168	4.1153	-	7470A	PPB	06 Aug 2019 14:33:40	U	SMPL
ICB60 - 1	ICB60	88	0.0606	-	7470A	PPB	06 Aug 2019 14:35:46	U	SMPL
CCV45 - 1	CCV45	19786	4.7368	-	7470A	PPB	06 Aug 2019 14:37:53	U	SMPL
CCB45 - 1	CCB45	61	0.0542	-	7470A	PPB	06 Aug 2019 14:39:59	U	SMPL
CRA - 1	CRA	782	0.2254	-	7470A	PPB	06 Aug 2019 14:42:06	U	SMPL
HighStd - 1	HighStd	41388	9.8651	-	7470A	PPB	06 Aug 2019 14:46:55	U	SMPL
ChkStd - 1	ChkStd	28832	6.8843	-	7470A	PPB	06 Aug 2019 14:49:01	U	SMPL
PB122019BL - 1	PBW	41	0.0494	-	7470A	PPB	06 Aug 2019 14:51:09	U	SMPL
PB122019BS - 1	LCSW	17144	4.1096	-	7470A	PPB	06 Aug 2019 14:55:47	U	SMPL
K4090-02 - 1	OR-02-080219-A	183	0.0832	-	7470A	PPB	06 Aug 2019 14:57:54	U	SMPL
K4090-02DUP - 1	OR-02-080219-ADUP	189	0.0846	-	7470A	PPB	06 Aug 2019 15:00:02	U	SMPL
K4090-02MS - 1	OR-02-080219-AMS	16667	3.9964	-	7470A	PPB	06 Aug 2019 15:02:10	U	SMPL
K4090-02MSD - 1	OR-02-080219-AMSD	16990	4.0731	-	7470A	PPB	06 Aug 2019 15:04:18	U	SMPL
K4090-04 - 1	OR-02-080219-B	268	0.1033	-	7470A	PPB	06 Aug 2019 15:06:25	U	SMPL
CCV46 - 1	CCV46	19477	4.6635	-	7470A	PPB	06 Aug 2019 15:08:32	U	SMPL
CCB46 - 1	CCB46	60	0.0540	-	7470A	PPB	06 Aug 2019 15:10:38	U	SMPL
K4093-02 - 1	PL-02-080119	354	0.1238	-	7470A	PPB	06 Aug 2019 15:12:45	U	SMPL
K4095-02 - 1	SU-02-080219-A	391	0.1325	-	7470A	PPB	06 Aug 2019 15:14:52	U	SMPL
K4095-04 - 1	SU-02-080219-B	268	0.1033	-	7470A	PPB	06 Aug 2019 15:16:59	U	SMPL
K4124-02 - 1	TP-15	217	0.0912	-	7470A	PPB	06 Aug 2019 15:19:05	U	SMPL
K4153-03 - 1	RT-4221	175	0.0813	-	7470A	PPB	06 Aug 2019 15:21:13	U	SMPL
K4155-01 - 1	TS-01_080219	303	0.1116	-	7470A	PPB	06 Aug 2019 15:23:20	U	SMPL
K4156-03 - 1	RT1494	338	0.1200	-	7470A	PPB	06 Aug 2019 15:25:28	U	SMPL
K4156-06 - 1	RT1873	326	0.1171	-	7470A	PPB	06 Aug 2019 15:27:36	U	SMPL
K4156-09 - 1	RT3380	272	0.1043	-	7470A	PPB	06 Aug 2019 15:29:45	U	SMPL
K4156-12 - 1	RT3456	258	0.1010	-	7470A	PPB	06 Aug 2019 15:31:51	U	SMPL
CCV47 - 1	CCV47	19406	4.6466	-	7470A	PPB	06 Aug 2019 15:33:58	U	SMPL
CCB47 - 1	CCB47	68	0.0559	-	7470A	PPB	06 Aug 2019 15:36:07	U	SMPL
K4156-15 - 1	RBR-200035	194	0.0858	-	7470A	PPB	06 Aug 2019 15:38:18	U	SMPL
PB122019TB - 1	PB121942TB	120	0.0682	-	7470A	PPB	06 Aug 2019 15:40:26	U	SMPL
K4090-02LX5 - 1		74	0.0573	-	7470A	PPB	06 Aug 2019 15:42:33	U	SMPL
K4090-02A - 1		16849	4.0396	-	7470A	PPB	06 Aug 2019 15:44:40	U	SMPL
CCV48 - 1	CCV48	19350	4.6333	-	7470A	PPB	06 Aug 2019 15:46:48	U	SMPL
CCB48 - 1	CCB48	40	0.0492	-	7470A	PPB	06 Aug 2019 15:48:55	U	SMPL