



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104445

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/06/2019	12:40
2	K4142-02		1	9	100.00	>212.0	>212.0	08/06/2019	13:10
3	K4142-02DUP		1	9	100.00	>212.0	>212.0	08/06/2019	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB104445

WorkList Name : FP8619 WorkList ID : 129456 Department : Wet-Chemistry Date : 08-06-2019 12:11:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/08/2019	Water	K4142-02	Flash Point	Cool 4 deg C	PSEG01	N21	399L-241L	08/02/2019	1010A

Date/Time 08/06/19 12:15
 Received by: KS
 Relinquished by: CP

Date/Time 08/06/19 13:45
 Received by: CP
 Relinquished by: KS