



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104448

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/06/2019	17:05
2	CAL2	1	Water	NA	NA	20.1	7.01	08/06/2019	17:06
3	CAL3	1	Water	NA	NA	20.1	10.02	08/06/2019	17:07
4	ICV	1	Water	NA	NA	20.1	7.00	08/06/2019	17:08
5	CCV1	1	Water	NA	NA	20.2	2.02	08/06/2019	17:10
6	K4195-05	1	Water	NA	NA	23.3	6.72	08/06/2019	17:11
7	K4195-05DUP	1	Water	NA	NA	23.4	6.73	08/06/2019	17:12
8	K4195-06	1	Water	NA	NA	23.6	6.63	08/06/2019	17:14
9	K4195-07	1	Water	NA	NA	23.8	6.78	08/06/2019	17:15
10	K4195-08	1	Water	NA	NA	24.2	8.17	08/06/2019	17:16
11	K4195-09	1	Water	NA	NA	24.1	8.33	08/06/2019	17:17
12	K4195-10	1	Water	NA	NA	23.7	6.78	08/06/2019	17:18
13	K4195-11	1	Water	NA	NA	23.7	6.82	08/06/2019	17:20
14	K4195-12	1	Water	NA	NA	23.0	6.97	08/06/2019	17:21
15	CCV2	1	Water	NA	NA	20.1	12.01	08/06/2019	17:22

23104448

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W K4158

Department : Wet-Chemistry

WorkList ID : 129472

Date : 08-06-2019 15:51:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2019	Water	K4195-05 ✓	pH	Cool 4 deg C	PSEG01	N23	30393	08/06/2019	9040C
08/06/2019	Water	K4195-06 ✓	pH	Cool 4 deg C	PSEG01	N23	30394	08/06/2019	9040C
08/06/2019	Water	K4195-07 ✓	pH	Cool 4 deg C	PSEG01	N23	30395	08/06/2019	9040C
08/06/2019	Water	K4195-08 ✓	pH	Cool 4 deg C	PSEG01	N23	30396	08/06/2019	9040C
08/06/2019	Water	K4195-09 ✓	pH	Cool 4 deg C	PSEG01	N23	30397	08/06/2019	9040C
08/06/2019	Water	K4195-10 ✓	pH	Cool 4 deg C	PSEG01	N23	30398	08/06/2019	9040C
08/06/2019	Water	K4195-11 ✓	pH	Cool 4 deg C	PSEG01	N23	30399	08/06/2019	9040C
08/06/2019	Water	K4195-12 ✓	pH	Cool 4 deg C	PSEG01	N23	30400	08/06/2019	9040C

Date/Time 08.06.19 16.00
 Received by: SA CD
 Relinquished by:

Date/Time 08.06.19 18.00
 Received by: SA
 Relinquished by: SA