



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104450

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/06/2019	17:05
2	CAL2	1	Water	NA	NA	20.1	7.01	08/06/2019	17:06
3	CAL3	1	Water	NA	NA	20.1	10.02	08/06/2019	17:07
4	ICV	1	Water	NA	NA	20.1	7.00	08/06/2019	17:08
5	CCV1	1	Water	NA	NA	20.2	2.02	08/06/2019	17:40
6	K4178-04	1	Solid	20.02	20	22.8	9.31	08/06/2019	17:41
7	K4191-01	1	Solid	20.03	20	21.7	7.23	08/06/2019	17:42
8	K4192-03	1	Solid	20.03	20	23.3	8.07	08/06/2019	17:43
9	K4192-05	1	Solid	20.04	20	23.1	7.92	08/06/2019	17:45
10	K4192-05DUP	1	Solid	20.03	20	23.2	7.93	08/06/2019	17:46
11	CCV2	1	Water	NA	NA	20.1	12.02	08/06/2019	17:48

VB 104430

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4192 WorkList ID : 129471 Department : Wet-Chemistry Date : 08-06-2019 15:51:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2019	Solid	K4178-04 - 7	Corrosivity	Cool 4 deg C	PSEG01	N23	370-371	08/06/2019	9045D
08/06/2019	Solid	K4191-01 - 7	Corrosivity	Cool 4 deg C	PSEG01	N23	VB-61627	08/06/2019	9045D
08/06/2019	Solid	K4192-03 - 7	Corrosivity	Cool 4 deg C	TULL02	P31	TRIPLE-BARREL-BC-WEST	08/06/2019	9045D
08/06/2019	Solid	K4192-05 - 7	Corrosivity	Cool 4 deg C	TULL02	P31	TRIPLE-BARREL-BC-EAST	08/06/2019	9045D

Date/Time 08/06/19 16:00
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 08/06/19 16:00
Received by: [Signature]
Relinquished by: [Signature]