


LB104459

7470A

INSTRUMENT ID : CU1

 μ Abs. :

41181



A= 0.0000e+000

B= 2.4296e-004

C= -1.5534e-002

Rho= 0.9998902

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	198	0.000	198				
0.2	0.200	0.192	-0.008	855	0.0 %	855				
2.5	2.500	2.414	-0.086	10000	0.0 %	10000				
5.0	5.000	5.094	0.094	21030	0.0 %	21030				
7.5	7.500	7.478	-0.022	30841	0.0 %	30841				
10.0	10.000	9.990	-0.010	41181	0.0 %	41181				

LB104459 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Aug 2019 08:53:40					
Sample ID	Extended ID	µ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type	
0.0 - 1	50-0	198	-	0.00007470A	PPB	07 Aug 2019 09:28:59	S	Std	
0.2 - 1	50-2	855	-	0.20007470A	PPB	07 Aug 2019 09:31:07	S	Std	
2.5 - 1	52-5	10000	-	2.50007470A	PPB	07 Aug 2019 09:33:16	S	Std	
5.0 - 1	55-0	21030	-	5.00007470A	PPB	07 Aug 2019 09:35:25	S	Std	
7.5 - 1	57-5	30841	-	7.50007470A	PPB	07 Aug 2019 09:37:33	S	Std	
10.0 - 1	510-0	41181	-	10.00007470A	PPB	07 Aug 2019 09:39:41	S	Std	
ICV61 - 1	ICV61	16603	4.0183	-7470A	PPB	07 Aug 2019 09:42:18	U	SMPL	
ICB61 - 1	ICB61	94	0.0073	-7470A	PPB	07 Aug 2019 09:46:59	U	SMPL	
CCV49 - 1	CCV49	20756	5.0273	-7470A	PPB	07 Aug 2019 09:49:05	U	SMPL	
CCB49 - 1	CCB49	50	-0.0034	-7470A	PPB	07 Aug 2019 09:51:11	U	SMPL	
CRA - 1	CRA	853	0.1917	-7470A	PPB	07 Aug 2019 09:53:17	U	SMPL	
HighStd - 1	HighStd	39377	9.5514	-7470A	PPB	07 Aug 2019 09:55:24	U	SMPL	
ChkStd - 1	ChkStd	27112	6.5716	-7470A	PPB	07 Aug 2019 09:57:31	U	SMPL	
PB122036BL - 1	PBW	253	0.0459	-7470A	PPB	07 Aug 2019 09:59:38	U	SMPL	
PB122036BS - 1	LCSW	16746	4.0531	-7470A	PPB	07 Aug 2019 10:01:46	U	SMPL	
K4129-15 - 1	2S-E1-E4-(0-1)COMP	888	0.2002	-7470A	PPB	07 Aug 2019 10:03:53	U	SMPL	
K4129-15DUP - 1	2S-E1-E4-(0-1)COMP DUP	886	0.1997	-7470A	PPB	07 Aug 2019 10:06:00	U	SMPL	
K4129-15MS - 1	2S-E1-E4-(0-1)COMPMS	17473	4.2297	-7470A	PPB	07 Aug 2019 10:08:08	U	SMPL	
K4129-15MSD - 1	2S-E1-E4-(0-1)COMPMSD	17818	4.3135	-7470A	PPB	07 Aug 2019 10:10:17	U	SMPL	
K4129-20 - 1	2S-E1-E4-(30)COMP	469	0.0984	-7470A	PPB	07 Aug 2019 10:12:24	U	SMPL	
CCV50 - 1	CCV50	21412	5.1867	-7470A	PPB	07 Aug 2019 10:14:30	U	SMPL	
CCB50 - 1	CCB50	80	0.0039	-7470A	PPB	07 Aug 2019 10:16:37	U	SMPL	
K4130-21 - 1	1S-F1-F4-COMP-(0-1)	174	0.0267	-7470A	PPB	07 Aug 2019 10:18:44	U	SMPL	
K4130-26 - 1	1S-F1-F4-COMP-(30)	277	0.0518	-7470A	PPB	07 Aug 2019 10:20:50	U	SMPL	
K4130-31 - 1	1S-G-(0-30)-COMP	361	0.0722	-7470A	PPB	07 Aug 2019 10:25:25	U	SMPL	
K4130-34 - 1	1S-L-(0-30)-COMP	443	0.0921	-7470A	PPB	07 Aug 2019 10:27:32	U	SMPL	
K4130-37 - 1	1S-M-(0-30)-COMP	599	0.1300	-7470A	PPB	07 Aug 2019 10:29:39	U	SMPL	
K4130-40 - 1	1S-H-(0-30)-COMP	646	0.1414	-7470A	PPB	07 Aug 2019 10:31:46	U	SMPL	
K4131-21 - 1	1S-D1-D4-COMP-(0-1)	83	0.0046	-7470A	PPB	07 Aug 2019 10:33:54	U	SMPL	
K4131-26 - 1	1S-D1-D4-COMP-(30)	169	0.0255	-7470A	PPB	07 Aug 2019 10:36:02	U	SMPL	
K4131-31 - 1	1S-E1-E4-COMP-(0-1)	302	0.0578	-7470A	PPB	07 Aug 2019 10:38:11	U	SMPL	
K4131-36 - 1	1S-E1-E4-COMP-(30)	409	0.0838	-7470A	PPB	07 Aug 2019 10:40:18	U	SMPL	
CCV51 - 1	CCV51	20749	5.0256	-7470A	PPB	07 Aug 2019 10:42:24	U	SMPL	
CCB51 - 1	CCB51	59	-0.0012	-7470A	PPB	07 Aug 2019 10:44:31	U	SMPL	
K4135-21 - 1	1S-A1-A4-COMP-(0-1)	532	0.1137	-7470A	PPB	07 Aug 2019 10:46:38	U	SMPL	
K4135-26 - 1	1S-A1-A4-COMP-(30)	665	0.1460	-7470A	PPB	07 Aug 2019 10:48:44	U	SMPL	
K4135-31 - 1	1S-B1-B4-COMP-(0-1)	149	0.0207	-7470A	PPB	07 Aug 2019 10:50:51	U	SMPL	
K4135-36 - 1	1S-B1-B4-COMP-(30)	158	0.0229	-7470A	PPB	07 Aug 2019 10:52:58	U	SMPL	
K4136-11 - 1	1S-C1-C4-COMP-(0-1)	199	0.0328	-7470A	PPB	07 Aug 2019 10:55:06	U	SMPL	
K4136-16 - 1	1S-C1-C4-COMP-(30)	193	0.0314	-7470A	PPB	07 Aug 2019 10:57:13	U	SMPL	
PB122037BL - 1	PBW	207	0.0348	-7470A	PPB	07 Aug 2019 10:59:22	U	SMPL	
PB122037BS - 1	LCSW	16648	4.0292	-7470A	PPB	07 Aug 2019 11:01:29	U	SMPL	
K2646-42 - 1	COMP-3	294	0.0559	-7470A	PPB	07 Aug 2019 11:03:39	U	SMPL	
K4142-02 - 1	399L-241L	288	0.0544	-7470A	PPB	07 Aug 2019 11:05:45	U	SMPL	
CCV52 - 1	CCV52	21031	5.0941	-7470A	PPB	07 Aug 2019 11:07:52	U	SMPL	
CCB52 - 1	CCB52	69	0.0012	-7470A	PPB	07 Aug 2019 11:10:00	U	SMPL	
K4142-02DUP - 1	399L-241LDUP	313	0.0605	-7470A	PPB	07 Aug 2019 11:12:07	U	SMPL	
K4142-02MS - 1	399L-241LMS	16472	3.9865	-7470A	PPB	07 Aug 2019 11:14:14	U	SMPL	
K4142-02MSD - 1	399L-241LMSD	16475	3.9872	-7470A	PPB	07 Aug 2019 11:16:21	U	SMPL	
PB122038BL - 1	PBW	180	0.0282	-7470A	PPB	07 Aug 2019 11:18:28	U	SMPL	
PB122038BS - 1	LCSW	16525	3.9994	-7470A	PPB	07 Aug 2019 11:20:36	U	SMPL	
K4152-21 - 1	2S-H1-H4-COMP-(0-1)	203	0.0338	-7470A	PPB	07 Aug 2019 11:22:44	U	SMPL	
K4152-26 - 1	2S-H1-H4-COMP-(30)	223	0.0386	-7470A	PPB	07 Aug 2019 11:24:52	U	SMPL	
K4152-31 - 1	2S-E1-E4-COMP-(0-1)	221	0.0382	-7470A	PPB	07 Aug 2019 11:27:00	U	SMPL	
K4152-36 - 1	2S-E1-E4-COMP-(30)	214	0.0365	-7470A	PPB	07 Aug 2019 11:29:09	U	SMPL	
K4154-11 - 1	PS-1	1116	0.2556	-7470A	PPB	07 Aug 2019 11:31:17	U	SMPL	
CCV53 - 1	CCV53	21030	5.0939	-7470A	PPB	07 Aug 2019 11:33:24	U	SMPL	
CCB53 - 1	CCB53	46	-0.0044	-7470A	PPB	07 Aug 2019 11:35:31	U	SMPL	
K4154-12 - 1	PS-2	903	0.2039	-7470A	PPB	07 Aug 2019 11:37:39	U	SMPL	
K4154-13 - 1	PS-3	511	0.1086	-7470A	PPB	07 Aug 2019 11:39:46	U	SMPL	
K4154-14 - 1	PS-4	237	0.0420	-7470A	PPB	07 Aug 2019 11:41:52	U	SMPL	
K4154-15 - 1	PS-5	231	0.0406	-7470A	PPB	07 Aug 2019 11:44:00	U	SMPL	
K4154-16 - 1	PS-6	257	0.0469	-7470A	PPB	07 Aug 2019 11:46:07	U	SMPL	
K4154-17 - 1	PS-7	905	0.2043	-7470A	PPB	07 Aug 2019 11:48:14	U	SMPL	
K4154-18 - 1	PS-8	698	0.1541	-7470A	PPB	07 Aug 2019 11:50:22	U	SMPL	
K4154-19 - 1	PS-9	605	0.1315	-7470A	PPB	07 Aug 2019 11:52:30	U	SMPL	
K4154-20 - 1	PS-10	230	0.0403	-7470A	PPB	07 Aug 2019 11:54:39	U	SMPL	
K4162-03 - 1	RT-3407	225	0.0391	-7470A	PPB	07 Aug 2019 11:56:47	U	SMPL	
CCV54 - 1	CCV54	20947	5.0737	-7470A	PPB	07 Aug 2019 11:58:56	U	SMPL	
CCB54 - 1	CCB54	65	0.0003	-7470A	PPB	07 Aug 2019 12:01:04	U	SMPL	
K4162-06 - 1	RBR-2000050	242	0.0433	-7470A	PPB	07 Aug 2019 12:03:12	U	SMPL	
K4164-03 - 1	RBR-2000030	975	0.2214	-7470A	PPB	07 Aug 2019 12:05:19	U	SMPL	
K4164-06 - 1	RBR-2000051	995	0.2262	-7470A	PPB	07 Aug 2019 12:07:27	U	SMPL	
K4164-06DUP - 1	RBR-2000051DUP	1007	0.2291	-7470A	PPB	07 Aug 2019 12:09:34	U	SMPL	
K4164-06MS - 1	RBR-2000051MS	17628	4.2673	-7470A	PPB	07 Aug 2019 12:11:42	U	SMPL	
K4164-06MSD - 1	RBR-2000051MSD	16716	4.0458	-7470A	PPB	07 Aug 2019 12:13:49	U	SMPL	
PB122039BL - 1	PBW	84	0.0049	-7470A	PPB	07 Aug 2019 12:15:58	U	SMPL	
PB122039BS - 1	LCSW	16766	4.0579	-7470A	PPB	07 Aug 2019 12:18:05	U	SMPL	
K4142-01 - 1	399L-241L	210	0.0355	-7470A	PPB	07 Aug 2019 12:20:14	U	SMPL	
K4142-01DUP - 1	399L-241LDUP	251	0.0454	-7470A	PPB	07 Aug 2019 12:22:22	U	SMPL	
CCV55 - 1	CCV55	21019	5.0912	-7470A	PPB	07 Aug 2019 12:24:29	U	SMPL	
CCB55 - 1	CCB55	66	0.0005	-7470A	PPB	07 Aug 2019 12:26:37	U	SMPL	
K4142-01MS - 1	399L-241LMS	15910	3.8499	-7470A	PPB	07 Aug 2019 12:28:44	U	SMPL	
K4142-01MSD - 1	399L-241LMSD	16172	3.9136	-7470A	PPB	07 Aug 2019 12:30:51	U	SMPL	
K4148-01 - 1	20190598-FIELD-BLANK	100	0.0088	-7470A	PPB	07 Aug 2019 12:32:58	U	SMPL	
K4148-02 - 1	20190599-SITE-WATER	217	0.0372	-7470A	PPB	07 Aug 2019 12:35:06	U	SMPL	
K4148-03 - 1	COMP-BH-1-ELU-20190610	205	0.0343	-7470A	PPB	07 Aug 2019 12:37:15	U	SMPL	
K4148-04 - 1	COMP-BH-2-ELU-20190611	188	0.0301	-7470A	PPB	07 Aug 2019 12:39:24	U	SMPL	
K4148-05 - 1	COMP-BH-3-ELU-20190612	158	0.0229	-7470A	PPB	07 Aug 2019 12:41:33	U	SMPL	
K4148-06 - 1	COMP-BH-4-ELU-20190613	168	0.0253	-7470A	PPB	07 Aug 2019 12:43:40	U	SMPL	
K4148-07 - 1	COMP-BH-6(159-02)-ELU-20190614	219	0.0377	-7470A	PPB	07 Aug 2019 12:45:49	U	SMPL	
PB122036TB - 1	PB121941TB	158	0.0229	-7470A	PPB	07 Aug 2019 12:47:57	U	SMPL	
CCV56 - 1	CCV56	20910	5.0647	-7470A	PPB	07 Aug 2019 12:50:05	U	SMPL	

LB104459 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Aug 2019 08:53:40					
Sample ID	Extended ID	μ Abs	Conc.	Stnd Conc	Method	Units	Date	Type	Type
CCB56 - 1	CCB56	51	-0.0031	-7470A	PPB		07 Aug 2019 12:52:12	U	SMPL
PB122037TB - 1	PB121952TB	149	0.0207	-7470A	PPB		07 Aug 2019 12:54:20	U	SMPL
PB122038TB - 1	PB121991TB	63	-0.0002	-7470A	PPB		07 Aug 2019 12:56:28	U	SMPL
K4129-15LX5 - 1		204	0.0340	-7470A	PPB		07 Aug 2019 12:58:36	U	SMPL
K4129-15A - 1		17395	4.2107	-7470A	PPB		07 Aug 2019 13:00:45	U	SMPL
K4142-02LX5 - 1		48	-0.0039	-7470A	PPB		07 Aug 2019 13:02:53	U	SMPL
K4142-02A - 1		16466	3.9850	-7470A	PPB		07 Aug 2019 13:05:01	U	SMPL
K4164-06LX5 - 1		202	0.0335	-7470A	PPB		07 Aug 2019 13:07:09	U	SMPL
K4164-06A - 1		17760	4.2994	-7470A	PPB		07 Aug 2019 13:09:16	U	SMPL
K4142-01LX5 - 1		41	-0.0056	-7470A	PPB		07 Aug 2019 13:11:25	U	SMPL
K4142-01A - 1		16039	3.8813	-7470A	PPB		07 Aug 2019 13:13:34	U	SMPL
CCV57 - 1	CCV57	20545	4.9761	-7470A	PPB		07 Aug 2019 13:15:42	U	SMPL
CCB57 - 1	CCB57	108	0.0107	-7470A	PPB		07 Aug 2019 13:17:50	U	SMPL