



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104464

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP77154
Turbidity Calibration std, 0NTU	WP77155
Turbidity Calibration std, 20NTU	WP77150
Turbidity Calibration std, 40NTU	WP77149
Turbidity Calibration std, 1NTU	WP77153
Turbidity Calibration std, 10NTU	WP77151
Turbidity Calibration std, 5NTU	WP77152
10 NTU Standard 500 ml	W2297

Intercept: 0.0519

Slope: 0.9604

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.289	08/07/2019	12:48
2	CAL2	1	1	0.996	08/07/2019	12:51
3	CAL3	5	1	4.870	08/07/2019	12:55
4	CAL4	10	1	9.650	08/07/2019	12:59
5	CAL5	20	1	18.800	08/07/2019	13:06
6	CAL6	40	1	38.700	08/07/2019	13:17



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB104464

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.490	9.827	08/07/2019	13:42
2	ICB		1	0.207	0.161	08/07/2019	13:48
3	CCV1	10	1	9.550	9.890	08/07/2019	13:54
4	CCB1		1	0.210	0.165	08/07/2019	13:59
5	LB104464BL		1	0.216	0.171	08/07/2019	14:05
6	K4205-01		1	10.900	11.295	08/07/2019	14:12
7	K4206-01		1	20.100	20.875	08/07/2019	14:17
8	K4206-01DUP		1	20.000	20.771	08/07/2019	14:24
9	CCV2	10	1	9.710	10.056	08/07/2019	14:30
10	CCB2		1	0.243	0.199	08/07/2019	14:34

WORKLIST(Hardcopy Internal Chain)

LB104464

WorkList Name : tur080719

WorkList ID : 129513

Department : Wet-Chemistry

Date : 08-07-2019 12:33:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/08/2019	Water	K4205-01	Turbidity	Cool 4 deg C	LIRO01	O33	UDS-4-26-80619	08/06/2019	SM2130 B
08/08/2019	Water	K4206-01	Turbidity	Cool 4 deg C	LIRO01	O33	UDS-4-40-80619	08/06/2019	SM2130 B

Date/Time 08/07/19 12:48
 Received by: 
 Relinquished by: _____

Date/Time 08/07/19 14:40
 Received by: _____
 Relinquished by: 