



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: chirag

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB104465

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP76219
COD calibration std. 100 ppm	WP76220
COD calibration std. 10 ppm	WP76221
COD calibration std. 150 ppm	WP76222
COD calibration std. 50 ppm	WP76223
COD calibration std. 0 ppm	WP76224
COD Digestion Vials Low Level 0-150Mg/L	W2528
COD ICV-LCS std, 50ppm	WP77085
COD CCV std, 50ppm	WP77086
COD Digestion Vials Low Level 0-150Mg/L	W2557

Temp In (C): 150	Date In: 08/07/2019	Time In: 09:02
Temp Out (C): 151	Date Out: 08/07/2019	Time Out: 11:02

Intercept: 0.2594

Slope: 1.0023

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/11/2019	12:00
2	CAL2	10	1	Water	11.000	07/11/2019	12:01
3	CAL3	50	1	Water	50.000	07/11/2019	12:02
4	CAL4	100	1	Water	100.000	07/11/2019	12:03
5	CAL5	150	1	Water	151.000	07/11/2019	12:04



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: chirag

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB104465

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.626	07/11/2019	12:24
2	ICB				1	Water	0.000	-0.259	07/11/2019	12:25
3	CCV1	50			1	Water	51.000	50.624	08/07/2019	12:35
4	CCB1				1	Water	0.000	-0.259	08/07/2019	12:36
5	LB104465BL				1	Water	0.000	-0.259	08/07/2019	12:37
6	LB104465BS	50			1	Water	51.000	50.624	08/07/2019	12:37
7	K4163-06				50	Water	43.000	42.643	08/07/2019	12:38
8	K4179-07				50	Water	63.000	62.597	08/07/2019	12:39
9	K4179-07DUP				50	Water	62.000	61.599	08/07/2019	12:40
10	K4179-07MS	50			50	Water	60.000	59.604	08/07/2019	12:42
11	K4179-07MSD	50			50	Water	60.000	59.604	08/07/2019	12:43
12	CCV2	50			1	Water	51.000	50.624	08/07/2019	12:44
13	CCB2				1	Water	0.000	-0.259	08/07/2019	12:45