



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104468

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/07/2019	12:25
2	K4195-01		1	9	100.00	>212.0	>212.0	08/07/2019	12:55
3	K4195-01DUP		1	9	100.00	>212.0	>212.0	08/07/2019	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104468

WorkList Name : FP8719 WorkList ID : 129509 Department : Wet-Chemistry Date : 08-07-2019 09:52:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2019	Water	K4195-01 B	Flash Point	Cool 4 deg C	PSEG01	N23	COMP-1	08/06/2019	1010A

Date/Time 08/07/19 12:00
 Received by: RS
 Relinquished by: CP

Date/Time 08/07/19 13:25
 Received by: CP
 Relinquished by: RS