



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/8/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 03:26
In Date: 08/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 08/08/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104470

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K4211-01	ETP21	1	1.14	9.94	8.79	86.9
K4211-02	ETP22	2	1.12	9.66	8.23	83.3
K4211-03	ETP23	3	1.12	9.62	8.28	84.2
K4211-04	ETP24	4	1.14	9.61	8.01	81.1
K4211-05	ETP25	5	1.14	9.84	8.3	82.3
K4211-06	ETP26	6	1.14	9.54	8.39	86.3
K4211-07	ETP27	7	1.13	9.9	8.63	85.5
K4211-08	ETP28	8	1.13	9.57	8.34	85.4
K4211-09	ETP29	9	1.13	9.7	8.81	89.6
K4211-10	ETP30	10	1.13	9.74	8.86	89.8
K4211-11	ETP41	11	1.14	9.77	8.28	82.7
K4211-12	ETP44	12	1.14	9.72	8.11	81.2
K4211-13	ETP45	13	1.13	9.84	8.24	81.6
K4211-14	ETP46	14	1.12	9.62	8.59	87.9
K4211-15	ETP47	15	1.12	9.92	8.72	86.4
K4211-16	ETP48	16	1.12	9.63	8.51	86.8
K4211-17	ETP49	17	1.13	9.95	8.89	88.0
K4211-18	ETP50	18	1.13	9.61	8.54	87.4
K4211-19	ETP51	19	1.13	9.86	8.61	85.7
K4211-20	ETP51MS	20	1.13	9.86	8.61	85.7
K4211-21	ETP51MSD	21	1.13	9.86	8.61	85.7
K4211-22	ETP52	22	1.13	9.61	8.14	82.7

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$