



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 8/8/2019

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 08/07/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 08/08/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB104476

| Lab ID   | Client Sample ID    | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|---------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K3613-01 | AU-05-080719-A      | 1     | 1.15              | 9.95                          | 8.57                              | 84.3    |
| K3613-03 | AU-05-080719-B      | 2     | 1.15              | 9.65                          | 8.5                               | 86.5    |
| K3613-05 | WILDWOOD-AVE-PILE   | 3     | 1.18              | 9.57                          | 8.98                              | 93.0    |
| K4063-06 | S3-02-A             | 4     | 1.19              | 9.78                          | 7.91                              | 78.2    |
| K4092-01 | EO-01-080719-A      | 5     | 1.15              | 9.92                          | 9.35                              | 93.5    |
| K4092-03 | EO-01-080719-B      | 6     | 1.17              | 9.6                           | 9.01                              | 93.0    |
| K4092-05 | EO-01-080719-C      | 7     | 1.13              | 9.89                          | 9.25                              | 92.7    |
| K4196-01 | WC-073119-SOIL      | 8     | 1.13              | 9.71                          | 6.05                              | 57.3    |
| K4197-01 | 390805731           | 9     | 1.15              | 9.58                          | 4.68                              | 41.9    |
| K4200-01 | TP-14               | 10    | 1.18              | 9.77                          | 8.68                              | 87.3    |
| K4200-03 | TP-14-1             | 11    | 1.14              | 9.88                          | 8.74                              | 87.0    |
| K4200-04 | TP-14-2             | 12    | 1.18              | 9.65                          | 8.6                               | 87.6    |
| K4203-01 | S8-01-S             | 13    | 1.17              | 9.55                          | 7.69                              | 77.8    |
| K4203-02 | S8-01-B             | 14    | 1.13              | 9.63                          | 8.79                              | 90.1    |
| K4203-03 | S8-02-S             | 15    | 1.1               | 9.5                           | 7.66                              | 78.1    |
| K4203-04 | S8-02-B             | 16    | 1.14              | 9.81                          | 8.89                              | 89.4    |
| K4203-05 | DUP-8               | 17    | 1.14              | 9.96                          | 8.98                              | 88.9    |
| K4204-06 | COMP-1              | 18    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K4207-01 | GS-1                | 19    | 1.14              | 9.72                          | 7.65                              | 75.9    |
| K4207-02 | GS-2                | 20    | 1.14              | 9.76                          | 9.29                              | 94.5    |
| K4208-01 | SB-03 (13-14)       | 21    | 1.13              | 9.67                          | 7.53                              | 74.9    |
| K4208-02 | SB-02 (0-2)         | 22    | 1.14              | 9.98                          | 9.05                              | 89.5    |
| K4208-03 | SB-02 (10-12)       | 23    | 1.14              | 9.65                          | 7.79                              | 78.1    |
| K4208-04 | SB-01 (7-9)         | 24    | 1.16              | 9.94                          | 7.82                              | 75.9    |
| K4219-01 | TP-A                | 25    | 1.15              | 9.67                          | 8.77                              | 89.4    |
| K4219-02 | TP-A                | 26    | 1.15              | 9.67                          | 8.77                              | 89.4    |
| K4219-03 | TP-B                | 27    | 1.15              | 9.98                          | 8.81                              | 86.7    |
| K4219-04 | TP-B                | 28    | 1.15              | 9.98                          | 8.81                              | 86.7    |
| K4219-05 | VNJ-239-217-225-230 | 29    | 1.15              | 9.98                          | 9.35                              | 92.9    |
| K4219-06 | VNJ-239-217-225-230 | 30    | 1.15              | 9.98                          | 9.35                              | 92.9    |
| K4219-07 | 355-35-76           | 31    | 1.16              | 9.81                          | 9.04                              | 91.1    |
| K4219-08 | 355-35-76           | 32    | 1.16              | 9.81                          | 9.04                              | 91.1    |
| K4223-01 | SP-1                | 33    | 1.14              | 9.98                          | 8.12                              | 79.0    |
| K4223-02 | SP-1                | 34    | 1.14              | 9.98                          | 8.12                              | 79.0    |
| K4223-03 | SP-2                | 35    | 1.14              | 9.93                          | 7.88                              | 76.7    |
| K4223-04 | SP-2                | 36    | 1.14              | 9.93                          | 7.88                              | 76.7    |
| K4223-05 | SP-3                | 37    | 1.15              | 9.51                          | 7.62                              | 77.4    |
| K4223-06 | SP-3                | 38    | 1.15              | 9.51                          | 7.62                              | 77.4    |
| K4223-07 | SP-4                | 39    | 1.16              | 9.8                           | 7.71                              | 75.8    |



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 8/8/2019

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 08/07/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 08/08/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB104476

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K4223-08 | SP-4             | 40    | 1.16              | 9.8                           | 7.71                              | 75.8    |
| K4223-09 | SP-5             | 41    | 1.16              | 9.71                          | 7.69                              | 76.4    |
| K4223-10 | SP-5             | 42    | 1.16              | 9.71                          | 7.69                              | 76.4    |
| K4223-11 | SP-6             | 43    | 1.14              | 9.76                          | 7.71                              | 76.2    |
| K4223-12 | SP-6             | 44    | 1.14              | 9.76                          | 7.71                              | 76.2    |
| K4223-13 | SP-7             | 45    | 1.15              | 9.74                          | 7.97                              | 79.4    |
| K4223-14 | SP-7             | 46    | 1.15              | 9.74                          | 7.97                              | 79.4    |
| K4223-15 | SP-8             | 47    | 1.15              | 9.96                          | 7.72                              | 74.6    |
| K4223-16 | SP-8             | 48    | 1.15              | 9.96                          | 7.72                              | 74.6    |
| K4223-17 | SP-9             | 49    | 1.14              | 9.97                          | 8.23                              | 80.3    |
| K4223-18 | SP-9             | 50    | 1.14              | 9.97                          | 8.23                              | 80.3    |
| K4223-19 | SP-10            | 51    | 1.15              | 9.95                          | 8.12                              | 79.2    |
| K4223-20 | SP-10            | 52    | 1.15              | 9.95                          | 8.12                              | 79.2    |
| K4223-21 | SP-11            | 53    | 1.15              | 9.72                          | 7.55                              | 74.7    |
| K4223-22 | SP-11            | 54    | 1.15              | 9.72                          | 7.55                              | 74.7    |
| K4223-23 | SP-12            | 55    | 1.15              | 9.71                          | 7.8                               | 77.7    |
| K4223-24 | SP-12            | 56    | 1.15              | 9.71                          | 7.8                               | 77.7    |
| K4223-25 | SP-13            | 57    | 1.14              | 9.84                          | 8.04                              | 79.3    |
| K4223-26 | SP-13            | 58    | 1.14              | 9.84                          | 8.04                              | 79.3    |
| K4223-27 | SP-14            | 59    | 1.16              | 9.71                          | 7.79                              | 77.5    |
| K4223-28 | SP-14            | 60    | 1.16              | 9.71                          | 7.79                              | 77.5    |
| K4223-29 | SP-15            | 61    | 1.15              | 9.95                          | 7.84                              | 76.0    |
| K4223-30 | SP-15            | 62    | 1.15              | 9.95                          | 7.84                              | 76.0    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

## PERCENT SOLID

Supervisor:  
Analyst: jignesh  
Date: 8/8/2019

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 08/07/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 08/08/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB104476

| Lab ID   | Client Sample ID    | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|---------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K3613-01 | AU-05-080719-A      | 1     | 1.15              | 9.95                          | 8.57                              | 84.3    |
| K3613-03 | AU-05-080719-B      | 2     | 1.15              | 9.65                          | 8.5                               | 86.5    |
| K3613-05 | WILDWOOD-AVE-PILE   | 3     | 1.18              | 9.57                          | 8.98                              | 93.0    |
| K4063-06 | S3-02-A             | 4     | 1.19              | 9.78                          | 7.91                              | 78.2    |
| K4092-01 | EO-01-080719-A      | 5     | 1.15              | 9.92                          | 9.35                              | 93.5    |
| K4092-03 | EO-01-080719-B      | 6     | 1.17              | 9.6                           | 9.01                              | 93.0    |
| K4092-05 | EO-01-080719-C      | 7     | 1.13              | 9.89                          | 9.25                              | 92.7    |
| K4196-01 | WC-073119-SOIL      | 8     | 1.13              | 9.71                          | 6.05                              | 57.3    |
| K4197-01 | 390805731           | 9     | 1.15              | 9.58                          | 4.68                              | 41.9    |
| K4200-01 | TP-14               | 10    | 1.18              | 9.77                          | 8.68                              | 87.3    |
| K4200-03 | TP-14-1             | 11    | 1.14              | 9.88                          | 8.74                              | 87.0    |
| K4200-04 | TP-14-2             | 12    | 1.18              | 9.65                          | 8.6                               | 87.6    |
| K4203-01 | S8-01-S             | 13    | 1.17              | 9.55                          | 7.69                              | 77.8    |
| K4203-02 | S8-01-B             | 14    | 1.13              | 9.63                          | 8.79                              | 90.1    |
| K4203-03 | S8-02-S             | 15    | 1.1               | 9.5                           | 7.66                              | 78.1    |
| K4203-04 | S8-02-B             | 16    | 1.14              | 9.81                          | 8.89                              | 89.4    |
| K4203-05 | DUP-8               | 17    | 1.14              | 9.96                          | 8.98                              | 88.9    |
| K4204-06 | COMP-1              | 18    | 1.00              | 2.00                          | 2.00                              | 100.0   |
| K4207-01 | GS-1                | 19    | 1.14              | 9.72                          | 7.65                              | 75.9    |
| K4207-02 | GS-2                | 20    | 1.14              | 9.76                          | 9.29                              | 94.5    |
| K4208-01 | SB-03 (13-14)       | 21    | 1.13              | 9.67                          | 7.53                              | 74.9    |
| K4208-02 | SB-02 (0-2)         | 22    | 1.14              | 9.98                          | 9.05                              | 89.5    |
| K4208-03 | SB-02 (10-12)       | 23    | 1.14              | 9.65                          | 7.79                              | 78.1    |
| K4208-04 | SB-01 (7-9)         | 24    | 1.16              | 9.94                          | 7.82                              | 75.9    |
| K4219-01 | TP-A                | 25    | 1.15              | 9.67                          | 8.77                              | 89.4    |
| K4219-02 | TP-A                | 26    | 1.15              | 9.67                          | 8.77                              | 89.4    |
| K4219-03 | TP-B                | 27    | 1.15              | 9.98                          | 8.81                              | 86.7    |
| K4219-04 | TP-B                | 28    | 1.15              | 9.98                          | 8.81                              | 86.7    |
| K4219-05 | VNJ-239-217-225-230 | 29    | 1.15              | 9.98                          | 9.35                              | 92.9    |
| K4219-06 | VNJ-239-217-225-230 | 30    | 1.15              | 9.98                          | 9.35                              | 92.9    |
| K4219-07 | 355-35-76           | 31    | 1.16              | 9.81                          | 9.04                              | 91.1    |
| K4219-08 | 355-35-76           | 32    | 1.16              | 9.81                          | 9.04                              | 91.1    |
| K4223-01 | SP-1                | 33    | 1.14              | 9.98                          | 8.12                              | 79.0    |
| K4223-02 | SP-1                | 34    | 1.14              | 9.98                          | 8.12                              | 79.0    |
| K4223-03 | SP-2                | 35    | 1.14              | 9.93                          | 7.88                              | 76.7    |
| K4223-04 | SP-2                | 36    | 1.14              | 9.93                          | 7.88                              | 76.7    |
| K4223-05 | SP-3                | 37    | 1.15              | 9.51                          | 7.62                              | 77.4    |
| K4223-06 | SP-3                | 38    | 1.15              | 9.51                          | 7.62                              | 77.4    |
| K4223-07 | SP-4                | 39    | 1.16              | 9.8                           | 7.71                              | 75.8    |

## PERCENT SOLID

Supervisor:  
Analyst: jignesh  
Date: 8/8/2019

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 08/07/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 08/08/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB104476

| Lab ID   | Client Sample ID | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| K4223-08 | SP-4             | 40    | 1.16              | 9.8                           | 7.71                              | 75.8    |
| K4223-09 | SP-5             | 41    | 1.16              | 9.71                          | 7.69                              | 76.4    |
| K4223-10 | SP-5             | 42    | 1.16              | 9.71                          | 7.69                              | 76.4    |
| K4223-11 | SP-6             | 43    | 1.14              | 9.76                          | 7.71                              | 76.2    |
| K4223-12 | SP-6             | 44    | 1.14              | 9.76                          | 7.71                              | 76.2    |
| K4223-13 | SP-7             | 45    | 1.15              | 9.74                          | 7.97                              | 79.4    |
| K4223-14 | SP-7             | 46    | 1.15              | 9.74                          | 7.97                              | 79.4    |
| K4223-15 | SP-8             | 47    | 1.15              | 9.96                          | 7.72                              | 74.6    |
| K4223-16 | SP-8             | 48    | 1.15              | 9.96                          | 7.72                              | 74.6    |
| K4223-17 | SP-9             | 49    | 1.14              | 9.97                          | 8.23                              | 80.3    |
| K4223-18 | SP-9             | 50    | 1.14              | 9.97                          | 8.23                              | 80.3    |
| K4223-19 | SP-10            | 51    | 1.15              | 9.95                          | 8.12                              | 79.2    |
| K4223-20 | SP-10            | 52    | 1.15              | 9.95                          | 8.12                              | 79.2    |
| K4223-21 | SP-11            | 53    | 1.15              | 9.72                          | 7.55                              | 74.7    |
| K4223-22 | SP-11            | 54    | 1.15              | 9.72                          | 7.55                              | 74.7    |
| K4223-23 | SP-12            | 55    | 1.15              | 9.71                          | 7.8                               | 77.7    |
| K4223-24 | SP-12            | 56    | 1.15              | 9.71                          | 7.8                               | 77.7    |
| K4223-25 | SP-13            | 57    | 1.14              | 9.84                          | 8.04                              | 79.3    |
| K4223-26 | SP-13            | 58    | 1.14              | 9.84                          | 8.04                              | 79.3    |
| K4223-27 | SP-14            | 59    | 1.16              | 9.71                          | 7.79                              | 77.5    |
| K4223-28 | SP-14            | 60    | 1.16              | 9.71                          | 7.79                              | 77.5    |
| K4223-29 | SP-15            | 61    | 1.15              | 9.95                          | 7.84                              | 76.0    |
| K4223-30 | SP-15            | 62    | 1.15              | 9.95                          | 7.84                              | 76.0    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

VB 104476

WorkList Name : %1-080719

WorkList ID : 129490

Department : Wet-Chemistry

Date : 08-07-2019 08:15:25

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample   | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-------------------|--------------|--------------|
| 08/12/2019 | Solid  | K3613-01 | Percent Solids | Cool 4 deg C | PSEG05   | N31              | AU-05-080719-A    | 08/07/2019   | Chemtech -SO |
| 08/12/2019 | Solid  | K3613-03 | Percent Solids | Cool 4 deg C | PSEG05   | N31              | AU-05-080719-B    | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K3613-05 | Percent Solids | Cool 4 deg C | PSEG05   | N31              | WILDWOOD-AVE-PILE | 08/07/2019   | Chemtech -SO |
| 08/05/2019 | Solid  | K4063-06 | Percent Solids | Cool 4 deg C | BSIG01   | O12              | S3-02-A           | 07/28/2019   | Chemtech -SO |
| 08/12/2019 | Solid  | K4092-01 | Percent Solids | Cool 4 deg C | PSEG05   | N12              | EO-01-080719-A    | 08/07/2019   | Chemtech -SO |
| 08/12/2019 | Solid  | K4092-03 | Percent Solids | Cool 4 deg C | PSEG05   | N12              | EO-01-080719-B    | 08/07/2019   | Chemtech -SO |
| 08/12/2019 | Solid  | K4092-05 | Percent Solids | Cool 4 deg C | PSEG05   | N12              | EO-01-080719-C    | 08/07/2019   | Chemtech -SO |
| 08/10/2019 | Solid  | K4196-01 | Percent Solids | Cool 4 deg C | PARS15   | P41              | WC-073119-SOIL    | 07/31/2019   | Chemtech -SO |
| 08/08/2019 | Solid  | K4197-01 | Percent Solids | Cool 4 deg C | GARD04   | O32              | 390805731         | 08/05/2019   | Chemtech -SO |
| 08/16/2019 | Solid  | K4200-01 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-14             | 08/06/2019   | Chemtech -SO |
| 08/16/2019 | Solid  | K4200-03 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-14-1           | 08/06/2019   | Chemtech -SO |
| 08/16/2019 | Solid  | K4200-04 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-14-2           | 08/06/2019   | Chemtech -SO |
| 08/08/2019 | Solid  | K4203-01 | Percent Solids | Cool 4 deg C | BSIG01   | O41              | S8-01-S           | 08/07/2019   | Chemtech -SO |
| 08/08/2019 | Solid  | K4203-02 | Percent Solids | Cool 4 deg C | BSIG01   | O41              | S8-01-B           | 08/07/2019   | Chemtech -SO |
| 08/08/2019 | Solid  | K4203-03 | Percent Solids | Cool 4 deg C | BSIG01   | O41              | S8-02-S           | 08/07/2019   | Chemtech -SO |
| 08/08/2019 | Solid  | K4203-04 | Percent Solids | Cool 4 deg C | BSIG01   | O41              | S8-02-B           | 08/07/2019   | Chemtech -SO |
| 08/08/2019 | Solid  | K4203-05 | Percent Solids | Cool 4 deg C | BSIG01   | O41              | DUP-8             | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4204-06 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | COMP-1            | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4207-01 | Percent Solids | Cool 4 deg C | URSC05   | P11              | GS-1              | 08/06/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4207-02 | Percent Solids | Cool 4 deg C | URSC05   | P11              | GS-2              | 08/06/2019   | Chemtech -SO |
| 08/16/2019 | Solid  | K4208-01 | Percent Solids | Cool 4 deg C | BART02   | O33              | SB-03(13-14)      | 08/06/2019   | Chemtech -SO |
| 08/16/2019 | Solid  | K4208-02 | Percent Solids | Cool 4 deg C | BART02   | O33              | SB-02(0-2)        | 08/06/2019   | Chemtech -SO |

Date/Time 08/07/19 16:00

Received by: J9

Relinquished by: C9

Date/Time 08/07/19 14:30

Received by: J9

Relinquished by: C9

# WORKLIST(Hardcopy Internal Chain)

✓ 9/10/19/16

Worklist Name : %1-080719

Worklist ID : 129490

Department : Wet-Chemistry

Date : 08-07-2019 08:15:25

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample     | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|---------------------|--------------|--------------|
| 08/16/2019 | Solid  | K4208-03 | Percent Solids | Cool 4 deg C | BART02   | O33              | SB-02(10-12)        | 08/06/2019   | Chemtech -SO |
| 08/16/2019 | Solid  | K4208-04 | Percent Solids | Cool 4 deg C | BART02   | O33              | SB-01(7-9)          | 08/06/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-01 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-A {              | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-02 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-A }              | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-03 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-B {              | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-04 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | TP-B }              | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-05 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | VNJ-239-217-225-230 | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-06 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | VNJ-239-217-225-230 | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-07 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | 355-35-76 {         | 08/07/2019   | Chemtech -SO |
| 08/14/2019 | Solid  | K4219-08 | Percent Solids | Cool 4 deg C | PSEG01   | N31              | 355-35-76 }         | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-01 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-1                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-02 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-1                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-03 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-2                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-04 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-2                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-05 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-3                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-06 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-3                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-07 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-4                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-08 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-4                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-09 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-5                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-10 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-5                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-11 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-6                | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-12 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-6                | 08/07/2019   | Chemtech -SO |

Date/Time 08/07/19 16:00

Received by: [Signature]

Relinquished by: [Signature]

Date/Time 08/07/19 14:30

Received by: [Signature]

Relinquished by: [Signature]



# WORKLIST(Hardcopy Internal Chain)

Y9 104476

Worklist Name : %1-080719

Worklist ID : 129490

Department : Wet-Chemistry

Date : 08-07-2019 08:15:25

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 08/17/2019 | Solid  | K4223-13 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-7            | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-14 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-7            | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-15 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-8            | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-16 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-8            | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-17 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-9            | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-18 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-9            | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-19 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-10           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-20 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-10           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-21 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-11           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-22 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-11           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-23 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-12           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-24 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-12           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-25 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-13           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-26 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-13           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-27 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-14           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-28 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-14           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-29 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-15           | 08/07/2019   | Chemtech -SO |
| 08/17/2019 | Solid  | K4223-30 | Percent Solids | Cool 4 deg C | DANN01   | O41              | SP-15           | 08/07/2019   | Chemtech -SO |

Date/Time 08/07/19 16100  
 Received by: JP  
 Relinquished by: CE

Date/Time 08/07/19 17430  
 Received by: CE  
 Relinquished by: JP