



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104477

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/07/2019	17:35
2	CAL2	1	Water	NA	NA	20.1	7.01	08/07/2019	17:36
3	CAL3	1	Water	NA	NA	20.1	10.02	08/07/2019	17:37
4	ICV	1	Water	NA	NA	20.1	6.99	08/07/2019	17:38
5	CCV1	1	Water	NA	NA	20.1	2.02	08/07/2019	17:39
6	K4196-01	1	Solid	20.02	20	21.4	10.75	08/07/2019	17:40
7	K4196-01DUP	1	Solid	20.03	20	21.5	10.76	08/07/2019	17:41
8	K4204-01	1	Solid	20.02	20	22.0	6.79	08/07/2019	17:42
9	K4204-02	1	Solid	20.03	20	21.8	6.85	08/07/2019	17:43
10	K4204-03	1	Solid	20.05	20	21.9	5.49	08/07/2019	17:45
11	K4204-04	1	Solid	20.02	20	21.9	4.62	08/07/2019	17:46
12	K4204-05	1	Solid	20.01	20	21.7	5.18	08/07/2019	17:48
13	K4204-06	1	Solid	20.04	20	21.9	4.79	08/07/2019	17:49
14	CCV2	1	Water	NA	NA	20.1	12.01	08/07/2019	17:50

RB 104477

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K4204

WorkList ID : 129540

Department : Wet-Chemistry

Date : 08-07-2019 17:16:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2019	Solid	K4196-01	pH	Cool 4 deg C	PARS15	P41	WC-073119-SOIL	07/31/2019	9045D
08/07/2019	Solid	K4204-01	pH	Cool 4 deg C	PSEG01	N31	13909	08/07/2019	9045D
08/07/2019	Solid	K4204-02	pH	Cool 4 deg C	PSEG01	N31	13912	08/07/2019	9045D
08/07/2019	Solid	K4204-03	pH	Cool 4 deg C	PSEG01	N31	13914	08/07/2019	9045D
08/07/2019	Solid	K4204-04	pH	Cool 4 deg C	PSEG01	N31	13913	08/07/2019	9045D
08/07/2019	Solid	K4204-05	pH	Cool 4 deg C	PSEG01	N31	30443	08/07/2019	9045D
08/07/2019	Solid	K4204-06	pH	Cool 4 deg C	PSEG01	N31	COMP-1	08/07/2019	9045D

Date/Time 08/07/19 17:20
 Received by: Ja
 Relinquished by: CB

Date/Time 08/07/19 18:35
 Received by: CB
 Relinquished by: JD