



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104478

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/07/2019	17:35
2	CAL2	1	Water	NA	NA	20.1	7.01	08/07/2019	17:36
3	CAL3	1	Water	NA	NA	20.1	10.02	08/07/2019	17:37
4	ICV	1	Water	NA	NA	20.1	6.99	08/07/2019	17:38
5	CCV1	1	Water	NA	NA	20.1	2.02	08/07/2019	18:00
6	K4204-08	1	Water	NA	NA	22.9	8.28	08/07/2019	18:01
7	K4204-08DUP	1	Water	NA	NA	22.8	8.29	08/07/2019	18:02
8	CCV2	1	Water	NA	NA	20.1	12.02	08/07/2019	18:04

VB 104478

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4204 WorkList ID : 129538 Department : Wet-Chemistry Date : 08-07-2019 17:15:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2019	Water	K4204-08 ✓	pH	Cool 4 deg C	PSEG01	N31	13967	08/07/2019	9040C

Date/Time 08/07/19 17120
Received by: Jg
Relinquished by: CL

Date/Time 08/07/19 18:35
Received by: CL
Relinquished by: Jg