



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104479

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/07/2019	17:35
2	CAL2	1	Water	NA	NA	20.1	7.01	08/07/2019	17:36
3	CAL3	1	Water	NA	NA	20.1	10.02	08/07/2019	17:37
4	ICV	1	Water	NA	NA	20.1	6.99	08/07/2019	17:38
5	CCV1	1	Water	NA	NA	20.1	2.02	08/07/2019	18:10
6	K3613-02	1	Solid	20.02	20	23.4	7.28	08/07/2019	18:11
7	K3613-02DUP	1	Solid	20.01	20	23.3	7.29	08/07/2019	18:12
8	K3613-04	1	Solid	20.03	20	23.0	7.77	08/07/2019	18:15
9	K3613-06	1	Solid	20.05	20	22.8	8.30	08/07/2019	18:16
10	K4092-02	1	Solid	20.04	20	23.5	8.94	08/07/2019	18:18
11	K4092-04	1	Solid	20.02	20	23.4	8.97	08/07/2019	18:20
12	K4092-06	1	Solid	20.01	20	22.8	8.80	08/07/2019	18:22
13	K4200-02	1	Solid	20.03	20	22.8	9.40	08/07/2019	18:25
14	K4219-02	1	Solid	20.02	20	23.8	7.32	08/07/2019	18:26
15	K4219-04	1	Solid	20.04	20	23.5	8.26	08/07/2019	18:30
16	CCV2	1	Water	NA	NA	20.1	12.02	08/07/2019	18:32
17	K4219-06	1	Solid	20.02	20	24.6	7.80	08/07/2019	18:33
18	K4219-08	1	Solid	20.01	20	24.4	7.75	08/07/2019	18:35
19	CCV3	1	Water	NA	NA	20.2	2.02	08/07/2019	18:36

RD 104479

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K36143 WorkList ID : 129536 Department : Wet-Chemistry Date : 08-07-2019 17:15:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2019	Solid	K3613-02- 	Corrosivity	Cool 4 deg C	PSEG05	N31	AU-05-080719-A	08/07/2019	9045D
08/07/2019	Solid	K3613-04- 	Corrosivity	Cool 4 deg C	PSEG05	N31	AU-05-080719-B	08/07/2019	9045D
08/07/2019	Solid	K3613-06- 	Corrosivity	Cool 4 deg C	PSEG05	N31	WILDWOOD-AVE-PILE	08/07/2019	9045D
08/07/2019	Solid	K4092-02- 	Corrosivity	Cool 4 deg C	PSEG05	N12	EO-01-080719-A	08/07/2019	9045D
08/07/2019	Solid	K4092-04- 	Corrosivity	Cool 4 deg C	PSEG05	N12	EO-01-080719-B	08/07/2019	9045D
08/07/2019	Solid	K4092-06- 	Corrosivity	Cool 4 deg C	PSEG05	N12	EO-01-080719-C	08/07/2019	9045D
08/06/2019	Solid	K4200-02- 	Corrosivity	Cool 4 deg C	PSEG01	N31	TP-14	08/06/2019	9045D
08/07/2019	Solid	K4219-02- 	Corrosivity	Cool 4 deg C	PSEG01	N31	TP-A	08/07/2019	9045D
08/07/2019	Solid	K4219-04- 	Corrosivity	Cool 4 deg C	PSEG01	N31	TP-B	08/07/2019	9045D
08/07/2019	Solid	K4219-06- 	Corrosivity	Cool 4 deg C	PSEG01	N31	VNJ-239-217-225-230	08/07/2019	9045D
08/07/2019	Solid	K4219-08- 	Corrosivity	Cool 4 deg C	PSEG01	N31	355-35-76	08/07/2019	9045D

Date/Time 08/07/19 17:120
 Received by: 
 Relinquished by: 

Date/Time 08/07/19 18:40
 Received by: 
 Relinquished by: 