



Analytical Summary Report

Analysis Method: 9040C

Analyst By : AMANDEEP

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104508

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	08/09/2019	08:15
2	CAL2	1	Water	NA	NA	20.1	6.99	08/09/2019	08:16
3	CAL3	1	Water	NA	NA	20.2	10.01	08/09/2019	08:17
4	ICV	1	Water	NA	NA	20.3	7.00	08/09/2019	08:18
5	CCV1	1	Water	NA	NA	20.1	2.01	08/09/2019	08:18
6	K4238-10	1	Water	NA	NA	22.2	7.20	08/09/2019	08:19
7	K4238-10DUP	1	Water	NA	NA	22.2	7.20	08/09/2019	08:19
8	CCV2	1	Water	NA	NA	20.2	12.01	08/09/2019	08:25

WORKLIST(Hardcopy Internal Chain)

60104508

WorkList Name : COOROSIVITY080919-2 WorkList ID : 129612 Department : Wet-Chemistry Date : 08-09-2019 07:57:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2019	Water	K4238-10	Corrosivity	Cool 4 deg C	PSEG01	N32	TP-17	08/07/2019	9040C

Date/Time 08/09/18 8:05
 Received by: AL
 Relinquished by: G

Date/Time 08/09/18 9:00
 Received by: Q
 Relinquished by: AL