



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : AMANDEEP

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104510

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	08/09/2019	08:15
2	CAL2	1	Water	NA	NA	20.1	6.99	08/09/2019	08:16
3	CAL3	1	Water	NA	NA	20.2	10.01	08/09/2019	08:17
4	ICV	1	Water	NA	NA	20.3	7.0	08/09/2019	08:18
5	CCV1	1	Water	NA	NA	20	1.99	08/09/2019	11:36
6	K4226-01	1	Water	NA	NA	18.9	6.60	08/09/2019	11:37
7	K4226-01DUP	1	Water	NA	NA	19.0	6.59	08/09/2019	11:38
8	CCV2	1	Water	NA	NA	20.2	11.98	08/09/2019	11:39

WORKLIST(Hardcopy Internal Chain)

LB104510

WorkList Name : PH080919-2

WorkList ID : 129632

Department : Wet-Chemistry

Date : 08-09-2019 11:27:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/08/2019	Water	K4226-01	pH	Cool 4 deg C	APT101	P11	WGGM-19-VW-TSF-032	08/08/2019	SM 4500-H B

Date/Time 08/09/19 11:35
 Received by: AK
 Relinquished by: Q

Date/Time 08/09/19 11:45
 Received by: Q
 Relinquished by: AK