


LB104521

7470A

INSTRUMENT ID: CV1

Linear ▼



A= 0.0000e+000

B= 2.4315e-004

C= 5.9372e-004

Rho= 0.9992622

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	80	0.000	80				
0.2	0.200	0.160	-0.040	654	0.0 %	654				
2.5	2.500	2.548	0.048	10477	0.0 %	10477				
5.0	5.000	4.857	-0.143	19971	0.0 %	19971				
7.5	7.500	7.769	0.269	31947	0.0 %	31947				
10.0	10.000	9.847	-0.153	40494	0.0 %	40494				

LB104521 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 09 Aug 2019 15:34:21					
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	80	-	0.0000	7470A	PPB	09 Aug 2019 15:56:26	S	Std
0.2 - 1	50.2	654	-	0.2000	7470A	PPB	09 Aug 2019 15:58:35	S	Std
2.5 - 1	52.5	10477	-	2.5000	7470A	PPB	09 Aug 2019 16:00:44	S	Std
5.0 - 1	55.0	19971	-	5.0000	7470A	PPB	09 Aug 2019 16:02:52	S	Std
7.5 - 1	57.5	31947	-	7.5000	7470A	PPB	09 Aug 2019 16:05:00	S	Std
10.0 - 1	510.0	40494	-	10.0000	7470A	PPB	09 Aug 2019 16:07:07	S	Std
ICV66 - 1	ICV66	16267	3.9560	-	7470A	PPB	09 Aug 2019 16:09:56	U	SMPL
ICB66 - 1	ICB66	-42	-0.0096	-	7470A	PPB	09 Aug 2019 16:12:02	U	SMPL
CCV86 - 1	CCV86	19755	4.8041	-	7470A	PPB	09 Aug 2019 16:14:09	U	SMPL
CCB86 - 1	CCB86	-39	-0.0089	-	7470A	PPB	09 Aug 2019 16:16:15	U	SMPL
CRA - 1	CRA	643	0.1569	-	7470A	PPB	09 Aug 2019 16:18:22	U	SMPL
HighStd - 1	HighStd	39387	9.5777	-	7470A	PPB	09 Aug 2019 16:20:28	U	SMPL
ChkStd - 1	ChkStd	28959	7.0421	-	7470A	PPB	09 Aug 2019 16:22:35	U	SMPL
PB122142BL - 1	PBW	-29	-0.0065	-	7470A	PPB	09 Aug 2019 16:24:43	U	SMPL
PB122142BS - 1	LCSW	16995	4.1330	-	7470A	PPB	09 Aug 2019 16:26:50	U	SMPL
K3613-02 - 1	AU-05-080719-A	-26	-0.0057	-	7470A	PPB	09 Aug 2019 16:28:57	U	SMPL
K3613-02DUP - 1	AU-05-080719-ADUP	5	0.0018	-	7470A	PPB	09 Aug 2019 16:31:05	U	SMPL
K3613-02MS - 1	AU-05-080719-AMS	16840	4.0953	-	7470A	PPB	09 Aug 2019 16:33:13	U	SMPL
K3613-02MSD - 1	AU-05-080719-AMSD	16877	4.1043	-	7470A	PPB	09 Aug 2019 16:35:22	U	SMPL
K3613-04 - 1	AU-05-080719-B	-29	-0.0065	-	7470A	PPB	09 Aug 2019 16:37:28	U	SMPL
CCV87 - 1	CCV87	19760	4.8053	-	7470A	PPB	09 Aug 2019 16:39:35	U	SMPL
CCB87 - 1	CCB87	-34	-0.0077	-	7470A	PPB	09 Aug 2019 16:41:42	U	SMPL
K3613-06 - 1	WILDWOOD-AVE-PILE	56	0.0142	-	7470A	PPB	09 Aug 2019 16:43:48	U	SMPL
K4092-02 - 1	EO-01-080719-A	40	0.0103	-	7470A	PPB	09 Aug 2019 16:45:55	U	SMPL
K4092-04 - 1	EO-01-080719-B	35	0.0091	-	7470A	PPB	09 Aug 2019 16:48:02	U	SMPL
K4092-06 - 1	EO-01-080719-C	42	0.0108	-	7470A	PPB	09 Aug 2019 16:50:08	U	SMPL
K4196-02 - 1	WC-073119-SOIL	36	0.0093	-	7470A	PPB	09 Aug 2019 16:52:16	U	SMPL
K4200-02 - 1	TP-14	12	0.0035	-	7470A	PPB	09 Aug 2019 16:54:23	U	SMPL
K4219-02 - 1	TP-A	29	0.0076	-	7470A	PPB	09 Aug 2019 16:56:31	U	SMPL
K4219-04 - 1	TP-B	55	0.0140	-	7470A	PPB	09 Aug 2019 16:58:39	U	SMPL
K4219-06 - 1	VNJ-239-217-225-230	52	0.0132	-	7470A	PPB	09 Aug 2019 17:00:47	U	SMPL
K4219-08 - 1	355-35-76	55	0.0140	-	7470A	PPB	09 Aug 2019 17:02:54	U	SMPL
CCV88 - 1	CCV88	19439	4.7273	-	7470A	PPB	09 Aug 2019 17:05:00	U	SMPL
CCB88 - 1	CCB88	-25	-0.0055	-	7470A	PPB	09 Aug 2019 17:07:07	U	SMPL
PB122143BL - 1	PBW	21	0.0057	-	7470A	PPB	09 Aug 2019 17:09:13	U	SMPL
PB122143BS - 1	LCSW	17240	4.1926	-	7470A	PPB	09 Aug 2019 17:11:20	U	SMPL
K4227-01 - 1	1S-I-(0-1)-COMP	-7	-0.0011	-	7470A	PPB	09 Aug 2019 17:13:29	U	SMPL
K4227-01DUP - 1	1S-I-(0-1)-COMP DUP	18	0.0050	-	7470A	PPB	09 Aug 2019 17:15:36	U	SMPL
K4227-01MS - 1	1S-I-(0-1)-COMPMS	16706	4.0627	-	7470A	PPB	09 Aug 2019 17:17:43	U	SMPL
K4227-01MSD - 1	1S-I-(0-1)-COMPMSD	17479	4.2507	-	7470A	PPB	09 Aug 2019 17:19:50	U	SMPL
K4227-04 - 1	1S-J-(0-1)-COMP	-32	-0.0072	-	7470A	PPB	09 Aug 2019 17:21:58	U	SMPL
K4227-07 - 1	1S-K-(0-1)-COMP	21	0.0057	-	7470A	PPB	09 Aug 2019 17:24:06	U	SMPL
K4227-10 - 1	1S-P-(0-1)-COMP	38	0.0098	-	7470A	PPB	09 Aug 2019 17:26:14	U	SMPL
K4227-13 - 1	1S-Q-(0-1)-COMP	43	0.0110	-	7470A	PPB	09 Aug 2019 17:28:21	U	SMPL
CCV89 - 1	CCV89	18922	4.6016	-	7470A	PPB	09 Aug 2019 17:30:28	U	SMPL
CCB89 - 1	CCB89	-20	-0.0043	-	7470A	PPB	09 Aug 2019 17:32:35	U	SMPL
K4227-16 - 1	1S-O-(0-1)-COMP	40	0.0103	-	7470A	PPB	09 Aug 2019 17:34:41	U	SMPL
K4227-19 - 1	1S-N-(0-1)-COMP	33	0.0086	-	7470A	PPB	09 Aug 2019 17:36:48	U	SMPL
K4228-22 - 1	2S-F1-F4-COMP-(0-1)	14	0.0040	-	7470A	PPB	09 Aug 2019 17:38:55	U	SMPL
K4228-27 - 1	2S-F1-F4-COMP-(1)	30	0.0079	-	7470A	PPB	09 Aug 2019 17:41:02	U	SMPL
K4228-32 - 1	2S-I1-I4-COMP-(0-1)	42	0.0108	-	7470A	PPB	09 Aug 2019 17:43:10	U	SMPL
K4228-37 - 1	2S-I1-I4-COMP-(1)	36	0.0093	-	7470A	PPB	09 Aug 2019 17:45:17	U	SMPL
K4241-02 - 1	TOTOWA-SOIL	35	0.0091	-	7470A	PPB	09 Aug 2019 17:47:25	U	SMPL
K4244-02 - 1	BERGEN-SP-1	47	0.0120	-	7470A	PPB	09 Aug 2019 17:49:33	U	SMPL
K4244-04 - 1	BERGEN-SP-2	23	0.0062	-	7470A	PPB	09 Aug 2019 17:51:42	U	SMPL
K4245-07 - 1	COMP-25	20	0.0055	-	7470A	PPB	09 Aug 2019 17:53:55	U	SMPL
CCV90 - 1	CCV90	18658	4.5374	-	7470A	PPB	09 Aug 2019 17:58:59	U	SMPL
CCB90 - 1	CCB90	-32	-0.0072	-	7470A	PPB	09 Aug 2019 18:01:05	U	SMPL
K4245-08 - 1	COMP-26	14	0.0040	-	7470A	PPB	09 Aug 2019 18:03:13	U	SMPL
K4245-09 - 1	COMP-27	21	0.0057	-	7470A	PPB	09 Aug 2019 18:05:20	U	SMPL
PB122144BL - 1	PBW	27	0.0072	-	7470A	PPB	09 Aug 2019 18:07:27	U	SMPL
PB122144BS - 1	LCSW	16976	4.1284	-	7470A	PPB	09 Aug 2019 18:09:34	U	SMPL
K4249-01 - 1	08-08-A	-3	-0.0001	-	7470A	PPB	09 Aug 2019 18:11:42	U	SMPL
K4249-01DUP - 1	08-08-ADUP	6	0.0021	-	7470A	PPB	09 Aug 2019 18:13:49	U	SMPL
K4249-01MS - 1	08-08-AMS	16149	3.9273	-	7470A	PPB	09 Aug 2019 18:15:57	U	SMPL
K4249-01MSD - 1	08-08-AMSD	16329	3.9711	-	7470A	PPB	09 Aug 2019 18:18:05	U	SMPL
K4249-02 - 1	08-08-B	-7	-0.0011	-	7470A	PPB	09 Aug 2019 18:20:14	U	SMPL
K4249-03 - 1	08-08-C	12	0.0035	-	7470A	PPB	09 Aug 2019 18:22:21	U	SMPL
CCV91 - 1	CCV91	19312	4.6964	-	7470A	PPB	09 Aug 2019 18:28:12	U	SMPL
CCB91 - 1	CCB91	-35	-0.0079	-	7470A	PPB	09 Aug 2019 18:30:19	U	SMPL
K4249-04 - 1	08-08-D	1	0.0008	-	7470A	PPB	09 Aug 2019 18:32:27	U	SMPL
K4249-05 - 1	08-08-E	9	0.0028	-	7470A	PPB	09 Aug 2019 18:34:34	U	SMPL
K4249-06 - 1	08-08-F	18	0.0050	-	7470A	PPB	09 Aug 2019 18:36:42	U	SMPL
K4249-07 - 1	08-08-G	24	0.0064	-	7470A	PPB	09 Aug 2019 18:38:49	U	SMPL
PB122145BL - 1	PBW	28	0.0074	-	7470A	PPB	09 Aug 2019 18:40:56	U	SMPL
PB122145BS - 1	LCSW	17640	4.2898	-	7470A	PPB	09 Aug 2019 18:43:04	U	SMPL
K4148-08 - 1	COMP-BH-1-ELU-20190610	71	0.0179	-	7470A	PPB	09 Aug 2019 18:45:12	U	SMPL
K4148-09 - 1	COMP-BH-2-ELU-20190611	123	0.0305	-	7470A	PPB	09 Aug 2019 18:47:20	U	SMPL
K4148-10 - 1	COMP-BH-2-ELU-20190612	107	0.0266	-	7470A	PPB	09 Aug 2019 18:49:29	U	SMPL
K4148-11 - 1	COMP-BH-2-ELU-20190613	82	0.0205	-	7470A	PPB	09 Aug 2019 18:51:37	U	SMPL
CCV92 - 1	CCV92	19781	4.8104	-	7470A	PPB	09 Aug 2019 18:53:45	U	SMPL
CCB92 - 1	CCB92	-29	-0.0065	-	7470A	PPB	09 Aug 2019 18:55:53	U	SMPL
K4148-12 - 1	COMP-BH-2-ELU-20190614	76	0.0191	-	7470A	PPB	09 Aug 2019 18:58:01	U	SMPL
K4238-09 - 1	TP-17	89265	21.7058	-	7470A	PPB	09 Aug 2019 19:06:22	U	SMPL
K4238-09DUP - 1	TP-17DUP	89200	21.6900	-	7470A	PPB	09 Aug 2019 19:08:30	U	SMPL
K4238-09MS - 1	TP-17MS	96971	23.5796	-	7470A	PPB	09 Aug 2019 19:10:39	U	SMPL
K4238-09MSD - 1	TP-17MSD	97546	23.7194	-	7470A	PPB	09 Aug 2019 19:12:47	U	SMPL
PB122142TB - 1	PB122060TB	-133	-0.0317	-	7470A	PPB	09 Aug 2019 19:14:56	U	SMPL
PB122143TB - 1	PB122099TB	-6	-0.0009	-	7470A	PPB	09 Aug 2019 19:17:05	U	SMPL
PB122144TB - 1	PB122116TB	21	0.0057	-	7470A	PPB	09 Aug 2019 19:19:13	U	SMPL
K3613-02LX5 - 1		14	0.0040	-	7470A	PPB	09 Aug 2019 19:21:22	U	SMPL
K3613-02A - 1		16782	4.0812	-	7470A	PPB	09 Aug 2019 19:23:29	U	SMPL
CCV93 - 1	CCV93	19576	4.7606	-	7470A	PPB	09 Aug 2019 19:25:37	U	SMPL

LB104521 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Aug 2019 15:34:21

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
CCB93 - 1	CCB93	-2	0.0001	-7470A	PPB	09 Aug 2019 19:27:44	U	SMPL
K4227-01LX5 - 1		0	0.0006	-7470A	PPB	09 Aug 2019 19:29:52	U	SMPL
K4227-01A - 1		16828	4.0924	-7470A	PPB	09 Aug 2019 19:31:59	U	SMPL
K4249-01LX5 - 1		-30	-0.0067	-7470A	PPB	09 Aug 2019 19:34:06	U	SMPL
K4249-01A - 1		16240	3.9494	-7470A	PPB	09 Aug 2019 19:36:13	U	SMPL
K4238-09LX5 - 1		16708	4.0632	-7470A	PPB	09 Aug 2019 19:38:22	U	SMPL
K4238-09A - 1		99053	24.0858	-7470A	PPB	09 Aug 2019 19:40:29	U	SMPL
K4238-09DLX5 - 1		16685	4.0576	-7470A	PPB	09 Aug 2019 19:42:37	U	SMPL
K4238-09DUPDLX5 - 1		16978	4.1289	-7470A	PPB	09 Aug 2019 19:44:46	U	SMPL
K4238-09MSDLX5 - 1		18457	4.4885	-7470A	PPB	09 Aug 2019 19:46:55	U	SMPL
K4238-09MSDDLX5 - 1		17327	4.2137	-7470A	PPB	09 Aug 2019 19:49:02	U	SMPL
CCV94 - 1	CCV94	19317	4.6976	-7470A	PPB	09 Aug 2019 19:51:10	U	SMPL
CCB94 - 1	CCB94	-19	-0.0040	-7470A	PPB	09 Aug 2019 19:53:18	U	SMPL
K4238-09DL2X25 - 1		3221	0.7838	-7470A	PPB	09 Aug 2019 19:55:26	U	SMPL
K4238-09ADLX5 - 1		18193	4.4243	-7470A	PPB	09 Aug 2019 19:57:34	U	SMPL
CCV95 - 1	CCV95	19092	4.6429	-7470A	PPB	09 Aug 2019 19:59:42	U	SMPL
CCB95 - 1	CCB95	-18	-0.0038	-7470A	PPB	09 Aug 2019 20:01:50	U	SMPL