



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104535

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/12/2019	08:55
2	K4238-10		1	8	100.00	>212.0	>212.0	08/12/2019	09:20
3	K4238-10DUP		1	8	100.00	>212.0	>212.0	08/12/2019	09:45
4	K4256-01		1	9	100.00	>212.0	>212.0	08/12/2019	10:10
5	K4256-03		1	8	100.00	>212.0	>212.0	08/12/2019	10:40
6	K4256-05		1	7	100.00	>212.0	>212.0	08/12/2019	11:05
7	K4277-14		1	7	100.00	>212.0	>212.0	08/12/2019	11:32
8	K4288-01		1	8	100.00	>212.0	>212.0	08/12/2019	11:58
9	K4288-15		1	9	100.00	>212.0	>212.0	08/12/2019	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8919 WorkList ID : 129652 Department : Wet-Chemistry Date : 08-09-2019 14:06:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2019	Water	K4238-10	Flash Point	Cool 4 deg C	PSEG01	N32	TP-17	08/07/2019	1010A
08/15/2019	Water	K4256-01	Flash Point	Cool 4 deg C	PSEG01	N32	30315	08/08/2019	1010A
08/15/2019	Water	K4256-03	Flash Point	Cool 4 deg C	PSEG01	N32	30334	08/08/2019	1010A
08/15/2019	Water	K4256-05	Flash Point	Cool 4 deg C	PSEG01	N32	COMP-2-30335-30428	08/08/2019	1010A
08/23/2019	Water	K4277-14	Flash Point	Cool 4 deg C	PSEG01	N33	TP-22	08/09/2019	1010A
08/16/2019	Water	K4288-01	Flash Point	Cool 4 deg C	PSEG01	N33	30361	08/09/2019	1010A
08/16/2019	Water	K4288-15	Flash Point	Cool 4 deg C	PSEG01	N33	COMP-5	08/09/2019	1010A

08/12/19
 Date/Time 08/12/19 08:30
 Received by: KS
 Relinquished by: CP

08/12/19 12:25
 Date/Time 08/12/19 12:25
 Received by: CP
 Relinquished by: KS