



Analytical Summary Report

Analysis Method: 9045D

Analyst By : AMANDEEP

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104540

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	08/12/2019	10:30
2	CAL2	1	Water	NA	NA	20.2	7.01	08/12/2019	10:38
3	CAL3	1	Water	NA	NA	20	10.02	08/12/2019	10:38
4	ICV	1	Water	NA	NA	20.3	6.98	08/12/2019	15:12
5	CCV1	1	Water	NA	NA	20.2	1.99	08/12/2019	11:12
6	K4288-03	1	Solid	20	20	21.4	5.29	08/12/2019	11:15
7	K4288-07	1	Solid	20	20	21.0	6.82	08/12/2019	11:23
8	K4288-08	1	Solid	20	20	21.0	8.55	08/12/2019	11:27
9	K4288-09	1	Solid	20	20	21.0	6.79	08/12/2019	11:30
10	K4288-10	1	Solid	20	20	20.8	6.89	08/12/2019	11:33
11	K4288-13	1	Solid	20	20	20.6	5.98	08/12/2019	11:35
12	K4288-14	1	Solid	20	20	20.9	6.00	08/12/2019	11:39
13	K4288-14DUP	1	Solid	20	20	20.8	5.58	08/12/2019	11:42
14	CCV2	1	Water	NA	NA	20.3	12.01	08/12/2019	11:45

WORKLIST(Hardcopy Internal Chain)

LB104540

WorkList Name : ph081219 WorkList ID : 129697 Department : Wet-Chemistry Date : 08-12-2019 10:06:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/09/2019	Solid	K4288-03	pH	Cool 4 deg C	PSEG01	N33	30359	08/09/2019	9041A
08/09/2019	Solid	K4288-07	pH	Cool 4 deg C	PSEG01	N33	30358	08/09/2019	9041A
08/09/2019	Solid	K4288-08	pH	Cool 4 deg C	PSEG01	N33	30360	08/09/2019	9041A
08/09/2019	Solid	K4288-09	pH	Cool 4 deg C	PSEG01	N33	30363	08/09/2019	9041A
08/09/2019	Solid	K4288-10	pH	Cool 4 deg C	PSEG01	N33	30364	08/09/2019	9041A
08/09/2019	Solid	K4288-13	pH	Cool 4 deg C	PSEG01	N33	30362	08/09/2019	9041A
08/09/2019	Solid	K4288-14	pH	Cool 4 deg C	PSEG01	N33	30355	08/09/2019	9041A

Date/Time 08-12-19 11:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 08-12-19 11:50
 Received by: [Signature]
 Relinquished by: [Signature]