



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : AMANDEEP

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104541

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	08/12/2019	10:30
2	CAL2	1	Water	NA	NA	20.2	7.01	08/12/2019	10:33
3	CAL3	1	Water	NA	NA	20.0	10.02	08/12/2019	10:38
4	ICV	1	Water	NA	NA	20.3	6.98	08/12/2019	10:45
5	CCV1	1	Water	NA	NA	20.2	2.02	08/12/2019	14:59
6	K4305-01	1	Water	NA	NA	21.3	4.10	08/12/2019	15:03
7	K4305-02	1	Water	NA	NA	21.4	7.87	08/12/2019	15:07
8	K4305-03	1	Water	NA	NA	21.0	5.50	08/12/2019	15:15
9	K4305-03DUP	1	Water	NA	NA	21.1	5.52	08/12/2019	15:18
10	CCV2	1	Water	NA	NA	20.5	11.98	08/12/2019	15:21

WORKLIST(Hardcopy Internal Chain)

LB104541

WorkList Name : PH081219-3 WorkList ID : 129725 Department : Wet-Chemistry Date : 08-12-2019 14:41:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2019	Water	K4305-01	A pH	Cool 4 deg C	STER17	O42	C-1	08/06/2019	SM 4500-H B
08/06/2019	Water	K4305-02	pH	Cool 4 deg C	STER17	O42	C-2	08/06/2019	SM 4500-H B
08/06/2019	Water	K4305-03	pH	Cool 4 deg C	STER17	O42	C-3	08/06/2019	SM 4500-H B

Date/Time 08-12-19 14:56
 Received by: ad
 Relinquished by: ad

Date/Time 08-12-19 15:30
 Received by: ad
 Relinquished by: ad