



Analytical Summary Report

Analysis Method: 9040C

Analyst By : AMANDEEP

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104542

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	08/12/2019	10:30
2	CAL2	1	Water	NA	NA	20.2	7.01	08/12/2019	10:33
3	CAL3	1	Water	NA	NA	20	10.02	08/12/2019	10:38
4	ICV	1	Water	NA	NA	20.3	6.98	08/12/2019	10:42
5	CCV1	1	Water	NA	NA	20.1	1.99	08/12/2019	10:45
6	K4277-13	1	Water	NA	NA	20.4	7.55	08/12/2019	10:49
7	K4288-01	1	Water	NA	NA	20.4	7.05	08/12/2019	10:58
8	K4288-17	1	Water	NA	NA	20.3	6.86	08/12/2019	11:02
9	K4288-18	1	Water	NA	NA	20.3	7.22	08/12/2019	11:06
10	K4288-18DUP	1	Water	NA	NA	20.4	7.21	08/12/2019	11:10
11	CCV2	1	Water	NA	NA	20.4	12.01	08/12/2019	11:18

WORKLIST(Hardcopy Internal Chain)

LB104542

WorkList Name : ph081219-2 WorkList ID : 129699 Department : Wet-Chemistry Date : 08-12-2019 10:07:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/09/2019	Water	K4277-13	pH	Cool 4 deg C	PSEG01	N33	TP-22	08/09/2019	9040C
08/09/2019	Water	K4288-01	— pH	Cool 4 deg C	PSEG01	N33	30361	08/09/2019	9040C
08/09/2019	Water	K4288-17	— pH	Cool 4 deg C	PSEG01	N33	30356	08/09/2019	9040C
08/09/2019	Water	K4288-18	— pH	Cool 4 deg C	PSEG01	N33	30357	08/09/2019	9040C

Date/Time 08-12-19 10:15
 Received by: AR
 Relinquished by: AR

Date/Time 08-12-19 11:20
 Received by: AR
 Relinquished by: AR