



PERCENT SOLID

Supervisor: apatel
Analyst: Dodley
Date: 8/13/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 20:14
In Date: 08/12/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 10:10
Out Date: 08/13/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104553

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K4085-01	HR-05-080919-A	1	1.13	9.98	9.12	90.3
K4085-03	HR-05-080919-B	2	1.13	9.63	8.85	90.8
K4085-05	HR-05-080919-C	3	1.14	9.6	8.84	91.0
K4090-05	OR-03-080919-A	4	1.14	9.74	8.87	89.9
K4090-07	OR-03-080919-B	5	1.14	9.87	9.01	90.1
K4095-05	SU-04-080919-A	6	1.13	9.97	9.34	92.9
K4095-07	SU-04-080919-B	7	1.13	9.68	8.83	90.1
K4280-01	VNJ-201	8	1.00	2.00	2.00	100.0
K4281-05	SVOC-GPC-BLANK	9	1.00	2.00	2.00	100.0
K4281-06	PEST-GPC-BLANK	10	1.00	2.00	2.00	100.0
K4281-07	PEST-GPC-BLANK-SPIKE	11	1.00	2.00	2.00	100.0
K4281-08	PCB-GPC-BLANK	12	1.00	2.00	2.00	100.0
K4281-09	PCB-GPC-BLANK-SPIKE	13	1.00	2.00	2.00	100.0
K4281-10	SVOC-GPC2-BLANK	14	1.00	2.00	2.00	100.0
K4281-11	PEST-GPC2-BLANK	15	1.00	2.00	2.00	100.0
K4281-12	PEST-GPC2-BLANK-SPIKE	16	1.00	2.00	2.00	100.0
K4281-13	PCB-GPC2-BLANK	17	1.00	2.00	2.00	100.0
K4281-14	PCB-GPC2-BLANK -SPIKE	18	1.00	2.00	2.00	100.0
K4284-01	CONCRETE-1	19	1.00	2.00	2.00	100.0
K4284-03	CONCRETE-2	20	1.00	2.00	2.00	100.0
K4284-05	CONCRETE-3	21	1.00	2.00	2.00	100.0
K4285-01	1908018037	22	1.00	2.00	2.00	100.0
K4288-03	30359	23	1.00	2.00	2.00	100.0
K4288-05	COMP-3	24	1.13	9.61	7.75	78.1
K4288-11	COMP-4	25	1.13	9.54	9.15	95.4
K4292-01	RO-COMP-WS	26	1.14	9.87	9.08	91.0
K4292-02	RO-COMP-WS	27	1.14	9.87	9.08	91.0
K4293-01	PE-1	28	1.15	9.64	8.39	85.3
K4293-02	PE-2	29	1.15	9.66	8.03	80.8
K4293-03	PE-3	30	1.15	9.83	8.15	80.6
K4293-04	PE-4	31	1.14	9.86	8.46	83.9
K4293-05	PE-5	32	1.14	9.8	8.2	81.5
K4294-01	RT-3425	33	1.13	9.71	9.56	98.3
K4294-02	RT-3425	34	1.15	9.92	9.73	97.8
K4294-04	RT-5124	35	1.15	9.63	8.72	89.3
K4294-05	RT-5124	36	1.14	9.62	8.38	85.4
K4294-07	RBR-200052	37	1.15	9.56	7.48	75.3
K4294-08	RBR-200052	38	1.13	9.59	7.48	75.1
K4294-10	RT-3860	39	1.13	9.98	9.87	98.8



PERCENT SOLID

Supervisor: apatel
Analyst: Dodley
Date: 8/13/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 20:14
In Date: 08/12/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 10:10
Out Date: 08/13/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104553

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
K4294-11	RT-3860	40	1.15	9.76	9.5	97.0
K4295-01	PX-01_080919	41	1.13	9.62	7.29	72.6
K4295-02	PX-02_080919	42	1.14	9.97	8.17	79.6
K4295-03	PX-03_080919	43	1.14	9.66	7.14	70.4
K4295-04	PX-04_080919	44	1.15	9.83	7.42	72.2
K4295-05	PX-05_080919	45	1.13	9.88	7.63	74.3
K4301-01	39080158	46	1.14	9.74	5.87	55.0
K4301-02	39080157	47	1.12	9.88	1.55	4.9

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$