

LB104555


ISM02.4

INSTRUMENT ID: CU1

Linear

 μ Abs.:

49211



A= 0.0000e+000

B= 2.0360e-004

C= 1.9731e-002

Rho= 0.9995999

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.084	0.084	317	0.000	317				
0.05	0.050									
0.20	0.200	0.218	0.018	972	0.0 %	972				
2.50	2.500	2.491	-0.009	12136	0.0 %	12136				
5.00	5.000	4.780	-0.220	23379	0.0 %	23379				
7.50	7.500	7.588	0.088	37174	0.0 %	37174				
10.0	10.000	10.039	0.039	49211	0.0 %	49211				

 $\frac{\%D}{10000}$

PS 08-13-15

LB104555
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	317	-		0 ISM02.4	PPB	8/13/2019 9:10	Std
	0.2 S0.2	972	-		0.2 ISM02.4	PPB	8/13/2019 9:13	Std
	2.5 S2.5	12136	-		2.5 ISM02.4	PPB	8/13/2019 9:15	Std
	5 S5.0	23379	-		5 ISM02.4	PPB	8/13/2019 9:17	Std
	7.5 S7.5	37174	-		7.5 ISM02.4	PPB	8/13/2019 9:19	Std
	10 S10.0	49211	-		10 ISM02.4	PPB	8/13/2019 9:22	Std
ICV	ICV	18952	3.8784 -		ISM02.4	PPB	8/13/2019 9:25	SMPL
ICB	ICB	320	0.0849 -		ISM02.4	PPB	8/13/2019 9:27	SMPL
CCV096	CCV096	23327	4.7692 -		ISM02.4	PPB	8/13/2019 9:29	SMPL
CCB096	CCB096	335	0.0879 -		ISM02.4	PPB	8/13/2019 9:31	SMPL
PB122222BL	PBS222	248	0.0702 -		ISM02.4	PPB	8/13/2019 9:34	SMPL
K4030-01	MBENZ9	35165	7.1794 -		ISM02.4	PPB	8/13/2019 9:36	SMPL
K4030-02	MBEP68	15235	3.1216 -		ISM02.4	PPB	8/13/2019 9:38	SMPL
K4030-03	MBEP69	17866	3.6573 -		ISM02.4	PPB	8/13/2019 9:40	SMPL
K4030-04	MBENQ6	20930	4.2811 -		ISM02.4	PPB	8/13/2019 9:43	SMPL
K4030-05	MBENQ7	35601	7.2682 -		ISM02.4	PPB	8/13/2019 9:45	SMPL
K4030-06	MBENY8	23201	4.7435 -		ISM02.4	PPB	8/13/2019 9:47	SMPL
K4030-07	MBENY9	26309	5.3763 -		ISM02.4	PPB	8/13/2019 9:50	SMPL
K4030-08	MBENZ0	54513	11.1187 -		ISM02.4	PPB	8/13/2019 9:52	SMPL
K4030-09	MBENZ1	11990	2.4609 -		ISM02.4	PPB	8/13/2019 9:54	SMPL
K4030-10	MBENZ2	19637	4.0179 -		ISM02.4	PPB	8/13/2019 9:56	SMPL
K4030-11	MBENZ2D	19545	3.9991 -		ISM02.4	PPB	8/13/2019 9:59	SMPL
K4030-12	MBENZ2S	29332	5.9918 -		ISM02.4	PPB	8/13/2019 10:01	SMPL
K4030-13	MBENZ3	20993	4.294 -		ISM02.4	PPB	8/13/2019 10:03	SMPL
K4030-14	MBENQ9	5825	1.2057 -		ISM02.4	PPB	8/13/2019 10:05	SMPL
K4030-15	MBENR0	1664	0.3585 -		ISM02.4	PPB	8/13/2019 10:08	SMPL
K4030-16	MBENR1	1610	0.3475 -		ISM02.4	PPB	8/13/2019 10:10	SMPL
K4030-17	MBENR4	26505	5.4162 -		ISM02.4	PPB	8/13/2019 10:12	SMPL
K4030-18	MBENR5	16220	3.3222 -		ISM02.4	PPB	8/13/2019 10:15	SMPL
K4030-19	MBEP71	1275	0.2793 -		ISM02.4	PPB	8/13/2019 10:17	SMPL
CCV097	CCV097	23503	4.805 -		ISM02.4	PPB	8/13/2019 10:19	SMPL
CCB097	CCB097	341	0.0892 -		ISM02.4	PPB	8/13/2019 10:21	SMPL
K4030-20	MBENW3	7963	1.641 -		ISM02.4	PPB	8/13/2019 10:24	SMPL
K4030-21	MBENW4	16040	3.2855 -		ISM02.4	PPB	8/13/2019 10:26	SMPL
PB122223BL	PBS223	323	0.0855 -		ISM02.4	PPB	8/13/2019 10:28	SMPL
K4031-01	MBENQ8	5091	1.0563 -		ISM02.4	PPB	8/13/2019 10:30	SMPL
K4031-02	MBENQ8D	5044	1.0467 -		ISM02.4	PPB	8/13/2019 10:33	SMPL
K4031-03	MBENQ8S	20191	4.1307 -		ISM02.4	PPB	8/13/2019 10:35	SMPL
K4031-04	MBENX5	16489	3.3769 -		ISM02.4	PPB	8/13/2019 10:37	SMPL
K4031-05	MBENX6	26994	5.5158 -		ISM02.4	PPB	8/13/2019 10:40	SMPL
K4031-06	MBENX7	17023	3.4857 -		ISM02.4	PPB	8/13/2019 10:42	SMPL
K4031-07	MBENX8	6769	1.3979 -		ISM02.4	PPB	8/13/2019 10:44	SMPL
K4031-08	MBENX9	7240	1.4938 -		ISM02.4	PPB	8/13/2019 10:46	SMPL
K4031-09	MBENY0	2245	0.4768 -		ISM02.4	PPB	8/13/2019 10:49	SMPL
K4031-10	MBENX4	7600	1.5671 -		ISM02.4	PPB	8/13/2019 10:51	SMPL
K4031-11	MBENY1	8458	1.7418 -		ISM02.4	PPB	8/13/2019 10:53	SMPL

LB104555

INSTRUMENT ID : CV1

K4031-12	MBENY2	7993	1.6471 -	ISM02.4	PPB	8/13/2019 10:56 SMPL
K4031-13	MBENY3	1413	0.3074 -	ISM02.4	PPB	8/13/2019 10:58 SMPL
K4031-14	MBENY4	11506	2.3624 -	ISM02.4	PPB	8/13/2019 11:00 SMPL
K4031-15	MBENY5	14915	3.0565 -	ISM02.4	PPB	8/13/2019 11:02 SMPL
K4031-16	MBENY6	7047	1.4545 -	ISM02.4	PPB	8/13/2019 11:05 SMPL
K4031-17	MBENT6	23780	4.8614 -	ISM02.4	PPB	8/13/2019 11:07 SMPL
CCV098	CCV098	23902	4.8862 -	ISM02.4	PPB	8/13/2019 11:09 SMPL
CCB098	CCB098	230	0.0666 -	ISM02.4	PPB	8/13/2019 11:11 SMPL
K4031-18	MBENT7	12936	2.6535 -	ISM02.4	PPB	8/13/2019 11:14 SMPL
K4031-19	MBENT8	16223	3.3228 -	ISM02.4	PPB	8/13/2019 11:16 SMPL
K4031-20	MBENT9	18746	3.8365 -	ISM02.4	PPB	8/13/2019 11:18 SMPL
K4031-21	MBENW0	31014	6.3343 -	ISM02.4	PPB	8/13/2019 11:21 SMPL
K4031-22	MBENW2	1305	0.2854 -	ISM02.4	PPB	8/13/2019 11:23 SMPL
K4030-08DLX2		26781	5.4724 -	ISM02.4	PPB	8/13/2019 11:32 SMPL
CCV099	CCV099	25100	5.1302 -	ISM02.4	PPB	8/13/2019 11:34 SMPL
CCB099	CCB099	147	0.0497 -	ISM02.4	PPB	8/13/2019 11:37 SMPL