


LB104558

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	71	0.000	71				
0.2	0.200	0.173	-0.027	808	0.0 %	808				
2.5	2.500	2.404	-0.096	10335	0.0 %	10335				
5.0	5.000	5.207	0.207	22304	0.0 %	22304				
7.5	7.500	7.506	0.006	32122	0.0 %	32122				
10.0	10.000	9.911	-0.089	42391	0.0 %	42391				

LB104558 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Aug 2019 14:33:44					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	71	-	0.00007470A	PPB		13 Aug 2019 14:56:58	S	Std
0.2 - 1	50.2	808	-	0.20007470A	PPB		13 Aug 2019 14:59:13	S	Std
2.5 - 1	52.5	10335	-	2.50007470A	PPB		13 Aug 2019 15:01:22	S	Std
5.0 - 1	55.0	22304	-	5.00007470A	PPB		13 Aug 2019 15:03:30	S	Std
7.5 - 1	57.5	32122	-	7.50007470A	PPB		13 Aug 2019 15:05:38	S	Std
10.0 - 1	510.0	42391	-	10.00007470A	PPB		13 Aug 2019 15:07:46	S	Std
ICV68 - 1	ICV68	17564	4.0967	-7470A	PPB		13 Aug 2019 15:10:23	U	SMPL
ICB68 - 1	ICB68	-15	-0.0200	-7470A	PPB		13 Aug 2019 15:12:29	U	SMPL
CCV01 - 1	CCV01	20886	4.8746	-7470A	PPB		13 Aug 2019 15:14:36	U	SMPL
CCB01 - 1	CCB01	-33	-0.0242	-7470A	PPB		13 Aug 2019 15:16:42	U	SMPL
CRA - 1	CRA	819	0.1753	-7470A	PPB		13 Aug 2019 15:18:48	U	SMPL
HighStd - 1	HighStd	41912	9.7986	-7470A	PPB		13 Aug 2019 15:20:55	U	SMPL
ChkStd - 1	ChkStd	29350	6.8568	-7470A	PPB		13 Aug 2019 15:23:02	U	SMPL
PB122234BL - 1	PBW	-15	-0.0200	-7470A	PPB		13 Aug 2019 15:25:09	U	SMPL
PB122234BS - 1	LCSW	17112	3.9908	-7470A	PPB		13 Aug 2019 15:27:17	U	SMPL
K4090-06 - 1	OR-03-080919-A	-3	-0.0172	-7470A	PPB		13 Aug 2019 15:29:24	U	SMPL
K4090-08 - 1	OR-03-080919-B	32	-0.0090	-7470A	PPB		13 Aug 2019 15:31:32	U	SMPL
K4229-02 - 1	2S-C1-C4-(0-1)-COMP	19	-0.0120	-7470A	PPB		13 Aug 2019 15:33:40	U	SMPL
K4229-02DUP - 1	2S-C1-C4-(0-1)-COMP	30	-0.0095	-7470A	PPB		13 Aug 2019 15:35:49	U	SMPL
K4229-02MS - 1	2S-C1-C4-(0-1)-COMPMS	16351	3.8126	-7470A	PPB		13 Aug 2019 15:37:56	U	SMPL
CCV02 - 1	CCV02	20669	4.8238	-7470A	PPB		13 Aug 2019 15:40:02	U	SMPL
CCB02 - 1	CCB02	-34	-0.0245	-7470A	PPB		13 Aug 2019 15:42:09	U	SMPL
K4229-02MSD - 1	2S-C1-C4-(0-1)-COMPMSD	15998	3.7300	-7470A	PPB		13 Aug 2019 15:44:16	U	SMPL
K4229-07 - 1	2S-C1-C4-(1)-COMP	-5	-0.0177	-7470A	PPB		13 Aug 2019 15:46:22	U	SMPL
K4229-12 - 1	2S-A1-A4-(0-1)-COMP	36	-0.0081	-7470A	PPB		13 Aug 2019 15:48:30	U	SMPL
K4229-17 - 1	2S-A1-A4-(1)-COMP	10	-0.0142	-7470A	PPB		13 Aug 2019 15:50:37	U	SMPL
K4230-03 - 1	2S-D1-D4-(0-1)-COMP	19	-0.0120	-7470A	PPB		13 Aug 2019 15:52:44	U	SMPL
K4230-08 - 1	2S-D1-D4-(1)-COMP	9	-0.0144	-7470A	PPB		13 Aug 2019 15:54:52	U	SMPL
K4230-12 - 1	2S-J-(0-1)-COMP	13	-0.0135	-7470A	PPB		13 Aug 2019 15:57:00	U	SMPL
K4230-15 - 1	REACTOR-3A-(0-3)-COMP	22	-0.0113	-7470A	PPB		13 Aug 2019 15:59:08	U	SMPL
K4230-18 - 1	REACTOR-3E-(0-3)-COMP	15	-0.0130	-7470A	PPB		13 Aug 2019 16:01:16	U	SMPL
K4231-02 - 1	REACTOR-3-F(0-3)	24	-0.0109	-7470A	PPB		13 Aug 2019 16:03:23	U	SMPL
CCV03 - 1	CCV03	21616	5.0456	-7470A	PPB		13 Aug 2019 16:05:30	U	SMPL
CCB03 - 1	CCB03	-14	-0.0198	-7470A	PPB		13 Aug 2019 16:07:37	U	SMPL
K4231-05 - 1	REACTOR-3-B(0-3)	9	-0.0144	-7470A	PPB		13 Aug 2019 16:09:44	U	SMPL
K4231-08 - 1	REACTOR-3-D(0-3)	17	-0.0125	-7470A	PPB		13 Aug 2019 16:11:50	U	SMPL
K4238-02 - 1	TP-17	13	-0.0135	-7470A	PPB		13 Aug 2019 16:13:57	U	SMPL
K4238-06 - 1	TP-18	29	-0.0097	-7470A	PPB		13 Aug 2019 16:16:05	U	SMPL
K4280-01 - 1	VNJ-201	19	-0.0120	-7470A	PPB		13 Aug 2019 16:18:13	U	SMPL
K4292-02 - 1	RO-COMP-WS	21	-0.0116	-7470A	PPB		13 Aug 2019 16:20:20	U	SMPL
PB122235BL - 1	PBW	45	-0.0060	-7470A	PPB		13 Aug 2019 16:22:28	U	SMPL
PB122235BS - 1	LCSW	17235	4.0196	-7470A	PPB		13 Aug 2019 16:24:36	U	SMPL
K4085-02 - 1	HR-05-080919-A	2	-0.0160	-7470A	PPB		13 Aug 2019 16:26:45	U	SMPL
K4085-04 - 1	HR-05-080919-B	21	-0.0116	-7470A	PPB		13 Aug 2019 16:28:52	U	SMPL
CCV04 - 1	CCV04	21802	5.0892	-7470A	PPB		13 Aug 2019 16:30:58	U	SMPL
CCB04 - 1	CCB04	-27	-0.0228	-7470A	PPB		13 Aug 2019 16:33:06	U	SMPL
K4085-06 - 1	HR-05-080919-C	33	-0.0088	-7470A	PPB		13 Aug 2019 16:35:13	U	SMPL
K4095-06 - 1	SU-04-080919-A	59	-0.0027	-7470A	PPB		13 Aug 2019 16:37:19	U	SMPL
K4095-08 - 1	SU-04-080919-B	9	-0.0144	-7470A	PPB		13 Aug 2019 16:39:26	U	SMPL
K4266-06 - 1	R1-R2(A1-A4)-(4.3)	41	-0.0069	-7470A	PPB		13 Aug 2019 16:41:33	U	SMPL
K4266-06DUP - 1	R1-R2(A1-A4)-(4.3)DUP	34	-0.0085	-7470A	PPB		13 Aug 2019 16:43:40	U	SMPL
K4266-06MS - 1	R1-R2(A1-A4)-(4.3)MS	16852	3.9299	-7470A	PPB		13 Aug 2019 16:45:48	U	SMPL
K4266-06MSD - 1	R1-R2(A1-A4)-(4.3)MSD	16630	3.8780	-7470A	PPB		13 Aug 2019 16:47:56	U	SMPL
K4266-21 - 1	R1-R2(A1-A4)-(0-1)	20	-0.0118	-7470A	PPB		13 Aug 2019 16:50:04	U	SMPL
K4266-23 - 1	R1-R2-C1-C4-(0-1)	9	-0.0144	-7470A	PPB		13 Aug 2019 16:52:12	U	SMPL
K4266-24 - 1	R1-R2-C1-C4-(4.3)	32	-0.0090	-7470A	PPB		13 Aug 2019 16:54:19	U	SMPL
CCV05 - 1	CCV05	21794	5.0873	-7470A	PPB		13 Aug 2019 16:56:27	U	SMPL
CCB05 - 1	CCB05	-9	-0.0186	-7470A	PPB		13 Aug 2019 16:58:34	U	SMPL
K4277-02 - 1	TP-19	36	-0.0081	-7470A	PPB		13 Aug 2019 17:00:41	U	SMPL
K4277-06 - 1	TP-20	59	-0.0027	-7470A	PPB		13 Aug 2019 17:02:48	U	SMPL
K4277-10 - 1	TP-22	52	-0.0043	-7470A	PPB		13 Aug 2019 17:04:55	U	SMPL
K4294-03 - 1	RT-3425	15	-0.0130	-7470A	PPB		13 Aug 2019 17:07:02	U	SMPL
K4294-06 - 1	RT-5124	23	-0.0111	-7470A	PPB		13 Aug 2019 17:09:09	U	SMPL
K4294-09 - 1	RBR-200052	18	-0.0123	-7470A	PPB		13 Aug 2019 17:11:16	U	SMPL
K4294-12 - 1	RT-3860	32	-0.0090	-7470A	PPB		13 Aug 2019 17:13:24	U	SMPL
PB122238BL - 1	PBW	8	-0.0146	-7470A	PPB		13 Aug 2019 17:15:32	U	SMPL
PB122238BS - 1	LCSW	17297	4.0342	-7470A	PPB		13 Aug 2019 17:17:41	U	SMPL
K4204-08 - 1	13967	3	-0.0158	-7470A	PPB		13 Aug 2019 17:19:48	U	SMPL
CCV06 - 1	CCV06	20800	4.8545	-7470A	PPB		13 Aug 2019 17:21:55	U	SMPL
CCB06 - 1	CCB06	-30	-0.0235	-7470A	PPB		13 Aug 2019 17:24:02	U	SMPL
K4238-10 - 1	TP-17	15	-0.0130	-7470A	PPB		13 Aug 2019 17:26:09	U	SMPL
K4277-14 - 1	TP-22	-5	-0.0177	-7470A	PPB		13 Aug 2019 17:28:17	U	SMPL
K4277-14DUP - 1	TP-22DUP	8	-0.0146	-7470A	PPB		13 Aug 2019 17:30:24	U	SMPL
K4277-14MS - 1	TP-22MS	15724	3.6658	-7470A	PPB		13 Aug 2019 17:32:31	U	SMPL
K4277-14MSD - 1	TP-22MSD	16002	3.7309	-7470A	PPB		13 Aug 2019 17:34:39	U	SMPL
PB122234TB - 1	PB122146TB	6	-0.0151	-7470A	PPB		13 Aug 2019 17:36:46	U	SMPL
PB122235TB - 1	PB122156TB	42	-0.0067	-7470A	PPB		13 Aug 2019 17:38:54	U	SMPL
PB122238TB - 1	PB122164TB	45	-0.0060	-7470A	PPB		13 Aug 2019 17:41:02	U	SMPL
K4229-02LX5 - 1		10	-0.0142	-7470A	PPB		13 Aug 2019 17:43:11	U	SMPL
K4229-02A - 1		16255	3.7901	-7470A	PPB		13 Aug 2019 17:45:18	U	SMPL
CCV07 - 1	CCV07	20483	4.7803	-7470A	PPB		13 Aug 2019 17:47:25	U	SMPL

LB104558 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Aug 2019 14:33:44					
Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
CCB07 - 1	CCB07	-35	-0.0247		-7470A	PPB	13 Aug 2019 17:49:32	U	SMPL
K4266-06LX5 - 1		-3	-0.0172		-7470A	PPB	13 Aug 2019 17:51:40	U	SMPL
K4266-06A - 1		16431	3.8314		-7470A	PPB	13 Aug 2019 17:53:47	U	SMPL
K4277-14LX5 - 1		-33	-0.0242		-7470A	PPB	13 Aug 2019 17:55:54	U	SMPL
K4277-14A - 1	CCV08	15994	3.7290		-7470A	PPB	13 Aug 2019 17:58:01	U	SMPL
CCV08 - 1		21436	5.0034		-7470A	PPB	13 Aug 2019 18:00:09	U	SMPL
CCB08 - 1	CCB08	-20	-0.0212		-7470A	PPB	13 Aug 2019 18:02:16	U	SMPL