



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104571

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/13/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	08/13/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	08/13/2019	17:02
4	ICV	1	Water	NA	NA	20.1	6.99	08/13/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	08/13/2019	17:06
6	K4329-04	1	Water	NA	NA	20.4	6.60	08/13/2019	17:07
7	K4329-04DUP	1	Water	NA	NA	20.5	6.61	08/13/2019	17:08
8	K4333-02	1	Water	NA	NA	21.4	11.56	08/13/2019	17:10
9	CCV2	1	Water	NA	NA	20.2	12.02	08/13/2019	17:46

2-B 10457

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4340 WorkList ID : 129777 Department : Wet-Chemistry Date : 08-13-2019 16:44:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2019	Water	K4329-04- <i>P</i>	Corrosivity	Cool 4 deg C	PSEG01	N42	WATER-PIT	08/13/2019	9040C
08/13/2019	Water	K4333-02- <i>P</i>	Corrosivity	Cool 4 deg C	PSEG01	N42	FRAC-TANK	08/13/2019	9040C

Date/Time 08.13.19 16:45
Received by: *JP*
Relinquished by: *CP*

Date/Time 08-13-19 18:30
Received by: *CP*
Relinquished by: *JP*