



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104573

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/13/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	08/13/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.02	08/13/2019	17:02
4	ICV	1	Water	NA	NA	20.1	6.99	08/13/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	08/13/2019	18:15
6	K4342-01	1	Solid	20.02	20	22.2	8.65	08/13/2019	18:16
7	K4342-01DUP	1	Solid	20.03	20	22.3	8.66	08/13/2019	18:17
8	CCV2	1	Water	NA	NA	20.1	12.02	08/13/2019	18:19

18104543

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S K4240 WorkList ID : 129775 Department : Wet-Chemistry Date : 08-13-2019 16:42:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/13/2019	Solid	K4342-01	pH	Cool 4 deg C	PSEG01	N42	CHRT-26872	08/13/2019	9045D

Date/Time 08.13.19 16:50
Received by: JJ
Relinquished by: ca

Date/Time 08.13.19 18:20
Received by: ca
Relinquished by: JJ