


ISM02.4

LB104580
INSTRUMENT ID : CV1



Linear

A= 0.0000e+000

B= 2.0820e-004

C= 1.6991e-002

Rho= 0.9998800

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.013	0.013	-21	0.000	-21				
0.05	0.050									
0.20	0.200	0.209	0.009	922	0.0 %	922				
2.50	2.500	2.484	-0.016	11850	0.0 %	11850				
5.00	5.000	4.930	-0.070	23596	0.0 %	23596				
7.50	7.500	7.610	0.110	36469	0.0 %	36469				
10.0	10.000	9.954	-0.046	47729	0.0 %	47729				

%D
—
5
-1
-1
0

RS 08.14.13

LB104580
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-21	-		0 ISM02.4	PPB	8/14/2019 7:55	Std
	0.2 S0.2	922	-		0.2 ISM02.4	PPB	8/14/2019 7:57	Std
	2.5 S2.5	11850	-		2.5 ISM02.4	PPB	8/14/2019 7:59	Std
	5 S5.0	23596	-		5 ISM02.4	PPB	8/14/2019 8:02	Std
	7.5 S7.5	36469	-		7.5 ISM02.4	PPB	8/14/2019 8:04	Std
	10 S10.0	47729	-		10 ISM02.4	PPB	8/14/2019 8:06	Std
ICV	ICV	18760	3.9229 -		ISM02.4	PPB	8/14/2019 8:14	SMPL
ICB	ICB	6	0.0182 -		ISM02.4	PPB	8/14/2019 8:16	SMPL
CCV009	CCV009	23786	4.9694 -		ISM02.4	PPB	8/14/2019 8:18	SMPL
CCB009	CCB009	-16	0.0137 -		ISM02.4	PPB	8/14/2019 8:20	SMPL
PB122224BL	PBS224	20	0.0212 -		ISM02.4	PPB	8/14/2019 8:23	SMPL
K3963-01	MBEP25	21350	4.4622 -		ISM02.4	PPB	8/14/2019 8:25	SMPL
K3963-02	MBEP26	25552	5.337 -		ISM02.4	PPB	8/14/2019 8:27	SMPL
K3963-03	MBEP29	14061	2.9446 -		ISM02.4	PPB	8/14/2019 8:30	SMPL
K3963-04	MBEP33	23108	4.8282 -		ISM02.4	PPB	8/14/2019 8:32	SMPL
K3963-05	MBEP27	16417	3.4351 -		ISM02.4	PPB	8/14/2019 8:34	SMPL
K3963-06	MBEP28	6963	1.4667 -		ISM02.4	PPB	8/14/2019 8:36	SMPL
K3963-07	MBEP30	27198	5.6797 -		ISM02.4	PPB	8/14/2019 8:39	SMPL
K3963-08	MBEP30D	27228	5.686 -		ISM02.4	PPB	8/14/2019 8:41	SMPL
K3963-09	MBEP30S	39698	8.2823 -		ISM02.4	PPB	8/14/2019 8:47	SMPL
K3963-10	MBEP03	27240	5.6885 -		ISM02.4	PPB	8/14/2019 8:49	SMPL
K3963-11	MBEP04	49035	10.2263 -		ISM02.4	PPB	8/14/2019 8:51	SMPL
K3963-12	MBEP05	20481	4.2812 -		ISM02.4	PPB	8/14/2019 8:53	SMPL
K3963-13	MBEP06	63158	13.1668 -		ISM02.4	PPB	8/14/2019 8:56	SMPL
K3963-14	MBEP07	6117	1.2906 -		ISM02.4	PPB	8/14/2019 8:58	SMPL
K3963-15	MBEP08	5968	1.2596 -		ISM02.4	PPB	8/14/2019 9:00	SMPL
K3963-16	MBEP09	2802	0.6004 -		ISM02.4	PPB	8/14/2019 9:02	SMPL
K3963-17	MBEP10	2775	0.5948 -		ISM02.4	PPB	8/14/2019 9:05	SMPL
K3963-18	MBENZ4	10256	2.1523 -		ISM02.4	PPB	8/14/2019 9:07	SMPL
K3963-19	MBENZ5	14011	2.9341 -		ISM02.4	PPB	8/14/2019 9:09	SMPL
CCV010	CCV010	24290	5.0743 -		ISM02.4	PPB	8/14/2019 9:12	SMPL
CCB010	CCB010	-1	0.0168 -		ISM02.4	PPB	8/14/2019 9:14	SMPL
K3963-20	MBENZ6	46160	9.6277 -		ISM02.4	PPB	8/14/2019 9:16	SMPL
K3963-21	MBENZ7	8904	1.8708 -		ISM02.4	PPB	8/14/2019 9:18	SMPL
K3963-22	MBENZ8	13298	2.7857 -		ISM02.4	PPB	8/14/2019 9:21	SMPL
K3963-11DLX2		24563	5.1311 -		ISM02.4	PPB	8/14/2019 9:23	SMPL
K3963-13DLX2		31144	6.5013 -		ISM02.4	PPB	8/14/2019 9:25	SMPL
CCV011	CCV011	23892	4.9914 -		ISM02.4	PPB	8/14/2019 9:27	SMPL
CCB011	CCB011	-32	0.0103 -		ISM02.4	PPB	8/14/2019 9:30	SMPL