



TOTAL VOLATILE SOLIDS 160.4

Run Number: LB104594

SUPERVISOR: apatel

ANALYST: AMANDEEP

BalanceID: WC SC-3

OvenID: WC Oven#1

TEMP1 IN:	104	08/13/2019 17:00	TEMP OUT:	105	08/14/2019 08:10
TEMP2 IN:	103	08/14/2019 09:20	TEMP OUT:	104	08/14/2019 10:25
TEMP3 IN:	550	08/14/2019 11:08	TEMP OUT:	550	08/14/2019 11:38
TEMP3 IN:	550	08/14/2019 12:05	TEMP OUT:	550	08/14/2019 13:30

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Empty Dish + Sample Weight (g)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result %
1	LB104594BL	53.5532	53.5532	53.5532	53.5532	53.5532	53.5532	53.5532	0	0
2	K4325-02	57.1838	57.1838	81.6061	79.9594	79.9594	79.7136	79.7136	0.2458	1.1
3	K4325-02DUP	60.8272	60.8272	84.209	82.5274	82.5274	82.2849	82.2849	0.2425	1.1

A = Sample Weight (g)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{F - E} \star 100$

WORKLIST(Hardcopy Internal Chain)

CB104598A

WorkList Name : TVS081319

WorkList ID : 129768

Department : Wet-Chemistry

Date : 08-13-2019 16:33:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2019	Solid	K4325-02	TVS	Cool 4 deg C	SEVE02	O43	20190612-LOC-14	08/12/2019	160.4

Date/Time 08-13-19 16:40
 Received by: JP
 Relinquished by: JP

Date/Time 08-13-19 17:15
 Received by: JP
 Relinquished by: JP