



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104595

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP77354
Turbidity Calibration std, 0NTU	WP77353
Turbidity Calibration std, 20NTU	WP77349
Turbidity Calibration std, 40NTU	WP77348
Turbidity Calibration std, 1NTU	WP77352
Turbidity Calibration std, 10NTU	WP77350
Turbidity Calibration std, 5NTU	WP77351
10 NTU Standard 500 ml	W2297

Intercept: 0.1509

Slope: 0.9237

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.198	08/14/2019	13:45
2	CAL2	1	1	0.986	08/14/2019	13:51
3	CAL3	5	1	4.800	08/14/2019	13:57
4	CAL4	10	1	9.620	08/14/2019	14:03
5	CAL5	20	1	18.300	08/14/2019	14:07
6	CAL6	40	1	37.200	08/14/2019	14:13



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104595

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.470	10.089	08/14/2019	14:16
2	ICB		1	0.198	0.051	08/14/2019	14:20
3	CCV1	10	1	9.520	10.143	08/14/2019	14:24
4	CCB1		1	0.201	0.054	08/14/2019	14:28
5	LB104595BL		1	0.214	0.068	08/14/2019	14:33
6	K4352-01		1	30.600	32.964	08/14/2019	14:37
7	K4352-01DUP		1	32.700	35.238	08/14/2019	14:41
8	CCV2	10	1	9.570	10.197	08/14/2019	14:45
9	CCB2		1	0.192	0.044	08/14/2019	14:47