



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104597

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/14/2019	09:15
2	K4329-04		1	8	100.00	>212.0	>212.0	08/14/2019	09:40
3	K4329-04DUP		1	8	100.00	>212.0	>212.0	08/14/2019	10:05
4	K4333-02		1	9	100.00	>212.0	>212.0	08/14/2019	10:35
5	K4338-01		1	7	100.00	>212.0	>212.0	08/14/2019	11:10
6	K4338-07		1	9	100.00	>212.0	>212.0	08/14/2019	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB104597

WorkList Name : FP81419 WorkList ID : 129817 Department : Wet-Chemistry Date : 08-14-2019 08:43:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method	
08/16/2019	Water	K4329-04	A	Flash Point	Cool 4 deg C	PSEG01	N42	WATER-PIT	08/13/2019	1010A
08/19/2019	Water	K4333-02	C	Flash Point	Cool 4 deg C	PSEG01	N42	FRAC-TANK	08/13/2019	1010A
08/20/2019	Water	K4338-01	B	Flash Point	Cool 4 deg C	PSEG01	N42	COMP-1	08/13/2019	1010A
08/20/2019	Water	K4338-07	B	Flash Point	Cool 4 deg C	PSEG01	N42	COMP-2	08/13/2019	1010A

Date/Time 08/14/19 08:45
 Received by: KS
 Relinquished by: JP

Date/Time 08/14/19 11:45
 Received by: JP
 Relinquished by: KS