

LB104601


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs.:

40079



A= 0.0000e+000

B= 2.4895e-004

C= -3.3792e-002

Rho= 0.9998592

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	299	0.000	299				
0.2	0.200	0.203	0.003	951	0.0 %	951				
2.5	2.500	2.411	-0.089	9820	0.0 %	9820				
5.0	5.000	5.001	0.001	20223	0.0 %	20223				
7.5	7.500	7.601	0.101	30669	0.0 %	30669				
10.0	10.000	9.944	-0.056	40079	0.0 %	40079				

LB104601 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 14 Aug 2019 14:54:25				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	299	-	0.00007470A	PPB	14 Aug 2019 15:05:49	S	Std
0.2 - 1	50-2	951	-	0.20007470A	PPB	14 Aug 2019 15:07:58	S	Std
2.5 - 1	52-5	9820	-	2.50007470A	PPB	14 Aug 2019 15:10:07	S	Std
5.0 - 1	55-0	20223	-	5.00007470A	PPB	14 Aug 2019 15:12:15	S	Std
7.5 - 1	57-5	30669	-	7.50007470A	PPB	14 Aug 2019 15:14:24	S	Std
10.0 - 1	510-0	40079	-	10.00007470A	PPB	14 Aug 2019 15:16:32	S	Std
ICV70 - 1	ICV70	16942	4.1839	-7470A	PPB	14 Aug 2019 15:21:23	U	SMPL
ICB70 - 1	ICB70	182	0.0115	-7470A	PPB	14 Aug 2019 15:23:29	U	SMPL
CCV19 - 1	CCV19	20447	5.0564	-7470A	PPB	14 Aug 2019 15:25:38	U	SMPL
CCB19 - 1	CCB19	176	0.0100	-7470A	PPB	14 Aug 2019 15:27:45	U	SMPL
CRA - 1	CRA	906	0.1918	-7470A	PPB	14 Aug 2019 15:29:52	U	SMPL
HighStd - 1	HighStd	39910	9.9017	-7470A	PPB	14 Aug 2019 15:31:58	U	SMPL
ChkStd - 1	ChkStd	28597	7.0853	-7470A	PPB	14 Aug 2019 15:34:05	U	SMPL
PB122281BL - 1	PBW	175	0.0098	-7470A	PPB	14 Aug 2019 15:36:12	U	SMPL
PB122281BS - 1	LCSW	16569	4.0910	-7470A	PPB	14 Aug 2019 15:38:19	U	SMPL
K4088-07 - 1	CL-01-081219	205	0.0172	-7470A	PPB	14 Aug 2019 15:40:27	U	SMPL
K4088-07DUP - 1	CL-01-081219DUP	248	0.0279	-7470A	PPB	14 Aug 2019 15:42:35	U	SMPL
K4088-07MS - 1	CL-01-081219MS	16732	4.1316	-7470A	PPB	14 Aug 2019 15:44:43	U	SMPL
K4088-07MSD - 1	CL-01-081219MSD	17068	4.2152	-7470A	PPB	14 Aug 2019 15:46:52	U	SMPL
K4089-02 - 1	OK-03-081219	784	0.1614	-7470A	PPB	14 Aug 2019 15:48:59	U	SMPL
CCV20 - 1	CCV20	20136	4.9790	-7470A	PPB	14 Aug 2019 15:51:06	U	SMPL
CCB20 - 1	CCB20	174	0.0095	-7470A	PPB	14 Aug 2019 15:53:13	U	SMPL
K4096-03 - 1	TR-04-081219	362	0.0563	-7470A	PPB	14 Aug 2019 15:55:20	U	SMPL
K4321-02 - 1	SP-1	256	0.0299	-7470A	PPB	14 Aug 2019 15:57:27	U	SMPL
K4321-04 - 1	SP-2	261	0.0312	-7470A	PPB	14 Aug 2019 15:59:33	U	SMPL
K4321-06 - 1	SP-3	234	0.0245	-7470A	PPB	14 Aug 2019 16:01:40	U	SMPL
K4322-01 - 1	SP-1	529	0.0979	-7470A	PPB	14 Aug 2019 16:03:48	U	SMPL
K4322-02 - 1	SP-2	1030	0.2226	-7470A	PPB	14 Aug 2019 16:05:56	U	SMPL
PB122281TB - 1	PB122205TB	242	0.0265	-7470A	PPB	14 Aug 2019 16:08:04	U	SMPL
K4088-07LX5 - 1		168	0.0080	-7470A	PPB	14 Aug 2019 16:10:12	U	SMPL
K4088-07A - 1		16795	4.1473	-7470A	PPB	14 Aug 2019 16:12:21	U	SMPL
CCV21 - 1	CCV21	20364	5.0358	-7470A	PPB	14 Aug 2019 16:14:28	U	SMPL
CCB21 - 1	CCB21	177	0.0103	-7470A	PPB	14 Aug 2019 16:16:34	U	SMPL