



### Analytical Summary Report

Analysis Method: 9040C  
Parameter: pH  
Run Number: LB104611

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/14/2019	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	08/14/2019	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	08/14/2019	17:17
4	ICV	1	Water	NA	NA	20.2	6.99	08/14/2019	17:19
5	CCV1	1	Water	NA	NA	20.1	2.01	08/14/2019	17:40
6	K4353-01	1	Water	NA	NA	25.4	4.84	08/14/2019	17:41
7	K4353-01DUP	1	Water	NA	NA	25.5	4.85	08/14/2019	17:42
8	CCV2	1	Water	NA	NA	20.1	12.01	08/14/2019	17:44

VB 104611

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4353 W

WorkList ID : 129839

Department : Wet-Chemistry

Date : 08-14-2019 16:45:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2019	Water	K4353-01	pH	Cool 4 deg C	PSEG01	N43	40080	08/14/2019	9040C

Date/Time 08.14.19 17:00  
 Received by: JO  
 Relinquished by: CB

Date/Time 08.14.19 18:00  
 Received by: CB  
 Relinquished by: JO