



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104634

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/15/2019	13:40
2	CAL2	1	Water	NA	NA	20.1	7.00	08/15/2019	13:41
3	CAL3	1	Water	NA	NA	20.1	10.02	08/15/2019	13:42
4	ICV	1	Water	NA	NA	20.1	6.99	08/15/2019	13:44
5	CCV1	1	Water	NA	NA	20.1	2.02	08/15/2019	13:46
6	K4367-01	1	Water	NA	NA	21.2	5.97	08/15/2019	13:48
7	K4367-01DUP	1	Water	NA	NA	21.3	5.98	08/15/2019	13:49
8	CCV2	1	Water	NA	NA	20.1	12.02	08/15/2019	13:50

13104634

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4367 WorkList ID : 129884 Department : Wet-Chemistry Date : 08-15-2019 13:34:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/14/2019	Water	K4367-01-13	pH	Cool 4 deg C	STER12	P51	CELL-POOL-WATER-3RD-Q	08/14/2019	SM 4500-H B

Date/Time 08.15.19 13:35
Received by: for
Relinquished by: ad

Date/Time 08.15.19 13:00
Received by: ad
Relinquished by: ad