



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104682

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/19/2019	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	08/19/2019	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	08/19/2019	16:02
4	ICV	1	Water	NA	NA	20.2	6.99	08/19/2019	16:04
5	CCV1	1	Water	NA	NA	20.2	2.02	08/19/2019	16:05
6	K4095-12	1	Solid	20.02	20	24.5	8.95	08/19/2019	16:06
7	K4096-06	1	Solid	20.03	20	24.8	7.62	08/19/2019	16:07
8	K4350-09	1	Solid	20.01	20	22.9	8.02	08/19/2019	16:08
9	K4350-10	1	Solid	20.04	20	22.1	8.01	08/19/2019	16:10
10	K4401-04	1	Solid	20.04	20	23.1	10.78	08/19/2019	16:11
11	K4401-04DUP	1	Solid	20.03	20	23.2	10.79	08/19/2019	16:12
12	CCV2	1	Water	NA	NA	20.2	12.02	08/19/2019	16:15