


LB104690

7471B

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 1.6870e-004

C= -1.3076e-001

Rho= 0.9999110

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.048	-0.048	493	0.000	493				
0.2	0.200	0.247	0.047	2239	0.0 %	2239				
2.5	2.500	2.517	0.017	15695	0.0 %	15695				
5.0	5.000	4.944	-0.056	30078	0.0 %	30078				
7.5	7.500	7.574	0.074	45669	0.0 %	45669				
10.0	10.000	9.966	-0.034	59850	0.0 %	59850				

LB104690 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Aug 2019 13:10:20

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	493	-	0.0000	7471B	PPB	20 Aug 2019 13:17:03	S	Std
0.2 - 1	50-2	2239	-	0.2000	7471B	PPB	20 Aug 2019 13:19:21	S	Std
2.5 - 1	52-5	15695	-	2.5000	7471B	PPB	20 Aug 2019 13:21:39	S	Std
5.0 - 1	55-8	30078	-	5.0000	7471B	PPB	20 Aug 2019 13:23:57	S	Std
7.5 - 1	57-5	45669	-	7.5000	7471B	PPB	20 Aug 2019 13:26:15	S	Std
10.0 - 1	510-0	59850	-	10.0000	7471B	PPB	20 Aug 2019 13:33:23	S	Std
ICV77 - 1	ICV77	23925	3.9055	-	7471B	PPB	20 Aug 2019 13:37:52	U	SMPL
ICB77 - 1	ICB77	293	-0.0813	-	7471B	PPB	20 Aug 2019 13:40:08	U	SMPL
CCV54 - 1	CCV54	30226	4.9685	-	7471B	PPB	20 Aug 2019 13:42:24	U	SMPL
CCB54 - 1	CCB54	302	-0.0798	-	7471B	PPB	20 Aug 2019 13:44:39	U	SMPL
CRA - 1	CRA	2205	0.2412	-	7471B	PPB	20 Aug 2019 13:46:56	U	SMPL
HighStd - 1	HighStd	59749	9.9492	-	7471B	PPB	20 Aug 2019 13:49:12	U	SMPL
ChkStd - 1	ChkStd	41896	6.9373	-	7471B	PPB	20 Aug 2019 13:51:28	U	SMPL
PB122416BL - 1	PBS	371	-0.0682	-	7471B	PPB	20 Aug 2019 13:53:45	U	SMPL
PB122416BS - 1	LCSS	21571	3.5084	-	7471B	PPB	20 Aug 2019 13:56:04	U	SMPL
K4388-16 - 1	T6-E1-E2-(0-2)	42699	7.0728	-	7471B	PPB	20 Aug 2019 13:58:22	U	SMPL
K4388-16DUP - 1	T6-E1-E2-(0-2)DUP	42731	7.0782	-	7471B	PPB	20 Aug 2019 14:00:39	U	SMPL
K4388-16MS - 1	T6-E1-E2-(0-2)MS	66605	11.1058	-	7471B	PPB	20 Aug 2019 14:09:54	U	SMPL
K4388-16MSD - 1	T6-E1-E2-(0-2)MSD	60770	10.1214	-	7471B	PPB	20 Aug 2019 14:12:12	U	SMPL
K4389-05 - 1	T4-F(0-2)	20885	3.3926	-	7471B	PPB	20 Aug 2019 14:14:30	U	SMPL
CCV55 - 1	CCV55	30146	4.9550	-	7471B	PPB	20 Aug 2019 14:16:47	U	SMPL
CCB55 - 1	CCB55	287	-0.0823	-	7471B	PPB	20 Aug 2019 14:19:04	U	SMPL
K4391-01 - 1	N-WALL-1	3084	0.3895	-	7471B	PPB	20 Aug 2019 14:21:20	U	SMPL
K4391-02 - 1	N-WALL-2	6492	0.9645	-	7471B	PPB	20 Aug 2019 14:23:37	U	SMPL
K4391-03 - 1	E-CORNER	3557	0.4693	-	7471B	PPB	20 Aug 2019 14:25:53	U	SMPL
K4391-04 - 1	BOTTOM-1	98995	16.5702	-	7471B	PPB	20 Aug 2019 14:28:09	U	SMPL
K4391-05 - 1	W-WALL	5801	0.8479	-	7471B	PPB	20 Aug 2019 14:30:26	U	SMPL
K4391-06 - 1	N-WALL-0	4082	0.5579	-	7471B	PPB	20 Aug 2019 14:32:45	U	SMPL
K4391-07 - 1	S-CORNER-1	6992	1.0488	-	7471B	PPB	20 Aug 2019 14:35:02	U	SMPL
K4391-08 - 1	S-CORNER-2	2689	0.3229	-	7471B	PPB	20 Aug 2019 14:37:20	U	SMPL
K4391-09 - 1	BOTTOM-2	3559	0.4697	-	7471B	PPB	20 Aug 2019 14:39:38	U	SMPL
K4401-03 - 1	TP-1359B	6499	0.9656	-	7471B	PPB	20 Aug 2019 14:41:54	U	SMPL
CCV56 - 1	CCV56	30388	4.9958	-	7471B	PPB	20 Aug 2019 14:44:10	U	SMPL
CCB56 - 1	CCB56	279	-0.0837	-	7471B	PPB	20 Aug 2019 14:46:27	U	SMPL
K4403-01 - 1	T1-H1-H4-(0-2.5)	36532	6.0324	-	7471B	PPB	20 Aug 2019 14:48:43	U	SMPL
K4403-10 - 1	T1-G-(0-2.5)	39058	6.4585	-	7471B	PPB	20 Aug 2019 14:50:58	U	SMPL
K4403-13 - 1	T1-I-(0-2.5)	36905	6.0953	-	7471B	PPB	20 Aug 2019 14:53:15	U	SMPL
K4403-16 - 1	T1-J-(0-2.5)	27454	4.5009	-	7471B	PPB	20 Aug 2019 14:55:32	U	SMPL
K4404-02 - 1	T1-L-(0-2.5)	40141	6.6412	-	7471B	PPB	20 Aug 2019 14:57:50	U	SMPL
PB122417BL - 1	PBS	365	-0.0692	-	7471B	PPB	20 Aug 2019 15:00:06	U	SMPL
PB122417BS - 1	LCSS	21762	3.5406	-	7471B	PPB	20 Aug 2019 15:02:24	U	SMPL
K4380-01 - 1	STOCKPILE-175-LANDERS-PIT	1799	0.1727	-	7471B	PPB	20 Aug 2019 15:04:41	U	SMPL
K4380-01DUP - 1	STOCKPILE-175-LANDERS-PITDUP	1781	0.1697	-	7471B	PPB	20 Aug 2019 15:07:00	U	SMPL
K4380-01MS - 1	STOCKPILE-175-LANDERS-PITMS	23505	3.8346	-	7471B	PPB	20 Aug 2019 15:09:16	U	SMPL
CCV57 - 1	CCV57	27951	4.5847	-	7471B	PPB	20 Aug 2019 15:11:32	U	SMPL
CCB57 - 1	CCB57	262	-0.0866	-	7471B	PPB	20 Aug 2019 15:13:49	U	SMPL
K4380-01MSD - 1	STOCKPILE-175-LANDERS-PITMSD	23755	3.8768	-	7471B	PPB	20 Aug 2019 15:16:08	U	SMPL
K4384-01 - 1	B-1(3.5-4)	20166	3.2713	-	7471B	PPB	20 Aug 2019 15:18:24	U	SMPL
K4384-02 - 1	B-2(1.5-2.5)	20897	3.3947	-	7471B	PPB	20 Aug 2019 15:20:41	U	SMPL
K4384-03 - 1	B-3(3-4)	51448	8.5488	-	7471B	PPB	20 Aug 2019 15:22:57	U	SMPL
K4384-04 - 1	B4(3-4)	37268	6.1565	-	7471B	PPB	20 Aug 2019 15:25:14	U	SMPL
K4385-01 - 1	T8-A1-A4-(0-2)	41122	6.8067	-	7471B	PPB	20 Aug 2019 15:27:33	U	SMPL
K4385-10 - 1	T8-B-(0-2)	26277	4.3023	-	7471B	PPB	20 Aug 2019 15:29:51	U	SMPL
K4385-13 - 1	T8-C-(0-2)	16435	2.6419	-	7471B	PPB	20 Aug 2019 15:32:09	U	SMPL
K4385-16 - 1	T8-D-(0-2)	12752	2.0206	-	7471B	PPB	20 Aug 2019 15:34:28	U	SMPL
K4385-19 - 1	T8-E1-E4-(0-2)	16969	2.7320	-	7471B	PPB	20 Aug 2019 15:36:44	U	SMPL
CCV58 - 1	CCV58	27719	4.5456	-	7471B	PPB	20 Aug 2019 15:39:00	U	SMPL
CCB58 - 1	CCB58	211	-0.0952	-	7471B	PPB	20 Aug 2019 15:41:17	U	SMPL
K4386-08 - 1	T8-F-(0-2)	18909	3.0593	-	7471B	PPB	20 Aug 2019 15:43:34	U	SMPL
K4386-11 - 1	T6-A1-A4-(0-2)	57600	9.5866	-	7471B	PPB	20 Aug 2019 15:45:51	U	SMPL
K4386-20 - 1	T6-B-(0-2)	35267	5.8189	-	7471B	PPB	20 Aug 2019 15:48:07	U	SMPL
K4387-03 - 1	T6-C-(0-2)	27898	4.5758	-	7471B	PPB	20 Aug 2019 15:50:25	U	SMPL
K4387-06 - 1	T6-D-(0-2)	13062	2.0729	-	7471B	PPB	20 Aug 2019 15:52:43	U	SMPL
K4387-09 - 1	T4-A1-A4-(0-2)	23374	3.8125	-	7471B	PPB	20 Aug 2019 15:55:00	U	SMPL
K4387-18 - 1	T4-B1-B4-(0-2)	31160	5.1261	-	7471B	PPB	20 Aug 2019 15:57:18	U	SMPL
K4388-07 - 1	T4-C-(0-2)	23943	3.9085	-	7471B	PPB	20 Aug 2019 15:59:35	U	SMPL
K4388-10 - 1	T4-D-(0-2)	19666	3.1870	-	7471B	PPB	20 Aug 2019 16:01:54	U	SMPL
K4388-13 - 1	T4-E-(0-2)	29687	4.8776	-	7471B	PPB	20 Aug 2019 16:04:11	U	SMPL
CCV59 - 1	CCV59	27880	4.5727	-	7471B	PPB	20 Aug 2019 16:06:28	U	SMPL
CCB59 - 1	CCB59	280	-0.0835	-	7471B	PPB	20 Aug 2019 16:08:46	U	SMPL
K4388-16LX5 - 1		8568	1.3147	-	7471B	PPB	20 Aug 2019 16:11:04	U	SMPL
K4388-16A - 1		66724	11.1259	-	7471B	PPB	20 Aug 2019 16:13:21	U	SMPL
K4380-01LX5 - 1		455	-0.0540	-	7471B	PPB	20 Aug 2019 16:15:37	U	SMPL
K4380-01A - 1		23325	3.8043	-	7471B	PPB	20 Aug 2019 16:17:56	U	SMPL
K4388-16DLX2 - 1		20398	3.3105	-	7471B	PPB	20 Aug 2019 16:20:12	U	SMPL
K4388-16DUPDLX2 - 1		20606	3.3456	-	7471B	PPB	20 Aug 2019 16:22:29	U	SMPL
K4388-16MSDLX2 - 1		30459	5.0078	-	7471B	PPB	20 Aug 2019 16:24:46	U	SMPL
K4388-16MSDDLX2 - 1		28843	4.7352	-	7471B	PPB	20 Aug 2019 16:27:04	U	SMPL
K4388-16DLX2X10 - 1		4264	0.5886	-	7471B	PPB	20 Aug 2019 16:29:23	U	SMPL
CCV60 - 1	CCV60	28396	4.6598	-	7471B	PPB	20 Aug 2019 16:33:57	U	SMPL
CCB60 - 1	CCB60	317	-0.0773	-	7471B	PPB	20 Aug 2019 16:36:15	U	SMPL
K4391-04DLX4 - 1		24521	4.0060	-	7471B	PPB	20 Aug 2019 16:38:33	U	SMPL
K4388-16ADLX2 - 1		30593	5.0304	-	7471B	PPB	20 Aug 2019 16:40:50	U	SMPL
CCV61 - 1	CCV61	28261	4.6370	-	7471B	PPB	20 Aug 2019 16:43:08	U	SMPL
CCB61 - 1	CCB61	278	-0.0839	-	7471B	PPB	20 Aug 2019 16:45:26	U	SMPL