



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104691

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/20/2019	12:25
2	CAL2	1	Water	NA	NA	20.1	7.00	08/20/2019	12:26
3	CAL3	1	Water	NA	NA	20.1	10.02	08/20/2019	12:27
4	ICV	1	Water	NA	NA	20.2	6.99	08/20/2019	12:30
5	CCV1	1	Water	NA	NA	20.2	2.02	08/20/2019	12:31
6	K4415-01	1	Water	NA	NA	18.8	6.41	08/20/2019	12:34
7	K4415-02	1	Water	NA	NA	17.5	10.00	08/20/2019	12:36
8	K4415-03	1	Water	NA	NA	17.0	8.04	08/20/2019	12:39
9	K4415-04	1	Water	NA	NA	16.2	9.64	08/20/2019	12:40
10	K4415-04DUP	1	Water	NA	NA	16.3	9.65	08/20/2019	12:42
11	CCV2	1	Water	NA	NA	20.1	12.01	08/20/2019	12:45

WORKLIST(Hardcopy Internal Chain)

Worklist Name :	ph k4415	Worklist ID :	130022	Department :	Wet-Chemistry	Date :	08-20-2019 12:14:03
------------------------	----------	----------------------	--------	---------------------	---------------	---------------	---------------------

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2019	Water	K4415-01	✓ pH	Cool 4 deg C	PSEG01	J11	30416	08/20/2019	9040C
08/20/2019	Water	K4415-02	✓ pH	Cool 4 deg C	PSEG01	J11	30417	08/20/2019	9040C
08/20/2019	Water	K4415-03	✓ pH	Cool 4 deg C	PSEG01	J11	30453	08/20/2019	9040C
08/20/2019	Water	K4415-04	✓ pH	Cool 4 deg C	PSEG01	J11	30496	08/20/2019	9040C

Date/Time 08/20/19 12:20
Received by: JH
Relinquished by: CS