



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB104699
BalanceID: WC SC-4

Analyst By : JIGNESH
Supervisor Review By : apatel
Slope : 98.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/20/2019	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	08/20/2019	16:11
3	CAL3	1	Water	NA	NA	20.1	10.04	08/20/2019	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	08/20/2019	16:14
5	CCV1	1	Water	NA	NA	20.2	2.01	08/20/2019	16:15
6	K4086-04	1	Solid	20.02	20	23.4	6.87	08/20/2019	16:16
7	K4402-02	1	Solid	20.01	20	21.9	11.65	08/20/2019	16:18
8	K4418-05	1	Solid	20.03	20	24.7	7.24	08/20/2019	16:20
9	K4421-15	1	Solid	20.04	20	22.8	10.57	08/20/2019	16:21
10	K4421-16	1	Solid	20.02	20	22.8	10.20	08/20/2019	16:24
11	K4421-17	1	Solid	20.03	20	22.6	8.20	08/20/2019	16:25
12	K4421-18	1	Solid	20.04	20	22.5	8.07	08/20/2019	16:27
13	K4421-19	1	Solid	20.03	20	23.2	8.81	08/20/2019	16:30
14	K4421-20	1	Solid	20.01	20	23.2	10.04	08/20/2019	16:32
15	K4421-21	1	Solid	20.04	20	23.3	7.94	08/20/2019	16:34
16	CCV2	1	Water	NA	NA	20.1	12.02	08/20/2019	16:35
17	K4421-21DUP	1	Solid	20.03	20	23.4	7.95	08/20/2019	16:26
18	CCV3	1	Water	NA	NA	20.1	2.02	08/20/2019	16:30

2B104699

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4421 WorkList ID : 130032 Department : Wet-Chemistry Date : 08-20-2019 15:15:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2019	Solid	K4086-04 - 13	Corrosivity	Cool 4 deg C	PSEG05	N12	NB-08-082019	08/20/2019	9045D
08/15/2019	Solid	K4402-02 - 13	Corrosivity	Cool 4 deg C	AREC01	P31	FRAC-TANK-MUD	08/15/2019	9045D
08/20/2019	Solid	K4418-05 - 13	Corrosivity	Cool 4 deg C	PSEG01	J11	CFVP-228L	08/20/2019	9045D
08/19/2019	Solid	K4421-15 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-7-COMP	08/19/2019	9045D
08/19/2019	Solid	K4421-16 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-6-COMP	08/19/2019	9045D
08/19/2019	Solid	K4421-17 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-5-COMP	08/19/2019	9045D
08/19/2019	Solid	K4421-18 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-4-COMP	08/19/2019	9045D
08/19/2019	Solid	K4421-19 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-3-COMP	08/19/2019	9045D
08/19/2019	Solid	K4421-20 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-2-COMP	08/19/2019	9045D
08/19/2019	Solid	K4421-21 - 13	Corrosivity	Cool 4 deg C	LIRO01	K51	SUNSET-PARK-SB-1-COMP	08/19/2019	9045D

Date/Time 08/20/19 16:00
 Received by: Jo
 Relinquished by: Jo

Date/Time 08/20/19 14:30
 Received by: Jo
 Relinquished by: Jo